

Appendix M. MAUDE Reports

Table M1. Reports on CGM Devices from the FDA MAUDE Database

Note: Key information is presented in this table, but detailed event descriptions are available at the [FDA MAUDE database](#).

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-------------------|-------------|--------------|--------------------|--------------|--|---|---|
| MW5161155 | 10/6/2024 4:00 | Injury | DEXCOM, INC. | 10/16/2024 4:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5160563 | | Death | DEXCOM, INC. | 10/3/2024 4:00 | QBJ | DEXCOM CONTINUOUS GLUCOSE MONITOR | | Insufficient Information |
| MW5160567 | 9/27/2024 4:00 | Injury | DEXCOM, INC. | 10/3/2024 4:00 | QBJ | DEXCOM G6 SENSOR | | Hyperglycemia; Hypoglycemia |
| MW5160467 | | Injury | DEXCOM, INC. | 10/2/2024 4:00 | QBJ | DEXCOM | | No Clinical Signs, Symptoms or Conditions |
| MW5160472 | | Injury | DEXCOM, INC. | 10/2/2024 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5160333 | | Death | DEXCOM, INC. | 9/30/2024 4:00 | QBJ | DEXCOM | | Insufficient Information |
| MW5160384 | | Malfunction | DEXCOM, INC. | 9/30/2024 4:00 | QBJ | DEXCOM CONTINUOUS GLUCOSE MONITOR SYSTEM | | Rash |
| MW5160420 | | Injury | DEXCOM, INC. | 9/30/2024 4:00 | QBJ | DEXCOM G6 | | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|----------------|--------------|-----------------------------------|---|---|
| MW5160 191 | | Injury | DEXCOM, INC. | 9/26/2024 4:00 | QBJ | DEXCOM CONTINUOUS GLUCOSE MONITOR | | Hyperglycemia |
| MW5160 099 | 9/19/2024 4:00 | Malfunction | DEXCOM, INC. | 9/25/2024 4:00 | QBJ | DEXCOM G6 SENSOR | | Insufficient Information |
| MW5160 103 | | Malfunction | DEXCOM, INC. | 9/25/2024 4:00 | QBJ | DEXCOM | | No Clinical Signs, Symptoms or Conditions |
| MW5160 104 | | Malfunction | DEXCOM, INC. | 9/25/2024 4:00 | QBJ | DEXCOM | | No Clinical Signs, Symptoms or Conditions |
| MW5160 013 | 9/16/2024 4:00 | Injury | DEXCOM, INC. | 9/24/2024 4:00 | QBJ | DEXCOM SENSOR | | Insufficient Information |
| MW5160 023 | 9/17/2024 4:00 | Injury | DEXCOM, INC. | 9/24/2024 4:00 | QBJ | DEXCOM CONTINUOUS GLUCOSE MONITOR | | No Clinical Signs, Symptoms or Conditions |
| MW5159 708 | 9/16/2024 4:00 | Injury | DEXCOM, INC. | 9/17/2024 4:00 | QBJ | DEXCOM BLOOD GLUCOSE MONITOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5159 446 | 9/5/2024 4:00 | Injury | DEXCOM, INC. | 9/9/2024 4:00 | QBJ | DEXCOM G6 SENSOR | High Test Results | Insufficient Information |
| MW5159 455 | 8/2/2024 4:00 | Injury | DEXCOM, INC. | 9/9/2024 4:00 | QBJ | DEXCOM G6 | | Hypoglycemia |
| MW5159 311 | 8/31/2024 4:00 | Injury | DEXCOM, INC. | 9/6/2024 4:00 | QBJ | DEXCOM SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5159 349 | 9/4/2024 4:00 | Injury | DEXCOM, INC. | 9/6/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|----------------|--------------|------------------|---|---|
| MW5159 056 | 8/26/2024 4:00 | Injury | DEXCOM, INC. | 8/29/2024 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5159 057 | 8/26/2024 4:00 | Injury | DEXCOM, INC. | 8/29/2024 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5159 058 | 8/26/2024 4:00 | Injury | DEXCOM, INC. | 8/29/2024 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5159 081 | 8/26/2024 4:00 | Injury | DEXCOM, INC. | 8/29/2024 4:00 | QBJ | DEXCOM G6 | High Readings; Device Sensing Problem | Hyperglycemia |
| MW5159 137 | 8/27/2024 4:00 | Malfunction | DEXCOM, INC. | 8/29/2024 4:00 | QBJ | DEXCOM G6 SENSOR | | Insufficient Information |
| MW5158 710 | | Injury | DEXCOM, INC. | 8/21/2024 4:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5158 575 | 8/14/2024 4:00 | Injury | DEXCOM, INC. | 8/16/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5158 302 | | Injury | DEXCOM, INC. | 8/9/2024 4:00 | QBJ | DEXCOM | | Insufficient Information |
| MW5158 257 | 8/5/2024 4:00 | Injury | DEXCOM, INC. | 8/7/2024 4:00 | QBJ | DEXCOM | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5157 864 | | Injury | DEXCOM, INC. | 7/30/2024 4:00 | QBJ | DEXCOM | | Hypoglycemia |
| MW5157 737 | 7/22/2024 4:00 | Injury | DEXCOM, INC. | 7/26/2024 4:00 | QBJ | DEXCOM G6 SENSOR | | Discomfort |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|----------------|--------------|------------------|---|---|
| MW5156 911 | 6/26/2024 4:00 | Malfunction | DEXCOM, INC. | 7/1/2024 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5156 505 | 6/17/2024 4:00 | Injury | DEXCOM, INC. | 6/21/2024 4:00 | QBJ | DEXCOM G6 | | Insufficient Information; No Clinical Signs, Symptoms or Conditions |
| MW5156 459 | | Injury | DEXCOM, INC. | 6/20/2024 4:00 | QBJ | DEXCOM | | Insufficient Information |
| MW5156 471 | | Injury | DEXCOM, INC. | 6/20/2024 4:00 | QBJ | DEXCOM | | Insufficient Information |
| MW5155 988 | | Injury | DEXCOM, INC. | 6/7/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |
| MW5155 768 | 6/3/2024 4:00 | Malfunction | DEXCOM, INC. | 6/4/2024 4:00 | QBJ | DEXCOM G6 | | Insufficient Information |
| MW5153 878 | 4/11/2024 4:00 | Malfunction | DEXCOM, INC. | 4/15/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |
| MW5153 513 | 4/3/2024 4:00 | Injury | DEXCOM, INC. | 4/3/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |
| MW5153 215 | 3/23/2024 4:00 | Injury | DEXCOM, INC. | 3/26/2024 4:00 | QBJ | DEXCOM SENSOR | High Readings | No Clinical Signs, Symptoms or Conditions |
| MW5153 027 | 3/18/2024 4:00 | Injury | DEXCOM, INC. | 3/20/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-------------------|-------------|--------------|-------------------|--------------|---|---|---|
| MW5153005 | 2/20/2024 5:00 | Malfunction | DEXCOM, INC. | 3/19/2024 4:00 | QBJ | DEXCOM G6 CONTINUEO US GLUCOSE MONITOR SYSTEM | | Foreign Body In Patient |
| MW5152444 | | Injury | DEXCOM, INC. | 3/6/2024 5:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5151801 | 2/11/2024 5:00 | Injury | DEXCOM, INC. | 2/21/2024 5:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5151825 | | Injury | DEXCOM, INC. | 2/21/2024 5:00 | QBJ | DEXCOM | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5151167 | 2/5/2024 5:00 | Injury | DEXCOM, INC. | 2/7/2024 5:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5150660 | 1/20/2024 5:00 | Injury | DEXCOM, INC. | 1/23/2024 5:00 | QBJ | DEXCOM G6 SENSOR | High Readings | No Clinical Signs, Symptoms or Conditions |
| MW5150681 | 1/21/2024 5:00 | Malfunction | DEXCOM, INC. | 1/23/2024 5:00 | QBJ | DEXCOM G6 TRANSMITTER | | No Clinical Signs, Symptoms or Conditions |
| MW5150490 | | Injury | DEXCOM, INC. | 1/17/2024 5:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|--------------------|-------------|--------------|--------------------|--------------|----------------------------|--|---|
| MW5150101 | | Injury | DEXCOM, INC. | 1/9/2024 5:00 | QBJ | DEXCOM G6 | High Readings; Device Sensing Problem | Hyperglycemia |
| MW5149414 | | Injury | DEXCOM, INC. | 12/19/2023 5:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5148679 | | Injury | DEXCOM, INC. | 12/1/2023 5:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | No Clinical Signs, Symptoms or Conditions |
| MW5148559 | | Injury | DEXCOM, INC. | 11/29/2023 5:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Failure to Sense | Hyperglycemia |
| MW5148569 | 11/23/2023 5:00 | Injury | DEXCOM, INC. | 11/29/2023 5:00 | QBJ | DEXCOM G6 CGM SENSOR | | Unspecified Infection; Rash |
| MW5148448 | | Injury | DEXCOM, INC. | 11/22/2023 5:00 | QBJ | DEXCOM G6 | High Readings; Calibration Problem | Insufficient Information |
| MW5148455 | 11/18/2023 5:00 | Malfunction | DEXCOM, INC. | 11/22/2023 5:00 | QBJ | DEXCOM G6 CGM SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5148368 | | Injury | DEXCOM, INC. | 11/21/2023 5:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5148145 | | Injury | DEXCOM, INC. | 11/14/2023 5:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5148148 | | Injury | DEXCOM, INC. | 11/14/2023 5:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-----------------|-------------|--------------|-----------------|--------------|-------------------------------------|---|---|
| MW5148 101 | | Injury | DEXCOM, INC. | 11/13/2023 5:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5148 022 | | Injury | DEXCOM, INC. | 11/9/2023 5:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5147 844 | | Injury | DEXCOM, INC. | 11/6/2023 5:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hypoglycemia |
| MW5147 801 | | Injury | DEXCOM, INC. | 11/3/2023 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Hyperglycemia |
| MW5147 758 | | Injury | DEXCOM, INC. | 11/2/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5147 654 | | Injury | DEXCOM, INC. | 11/1/2023 4:00 | QBJ | DEXCOM G6 | | Insufficient Information |
| MW5147 085 | 10/12/2023 4:00 | Malfunction | DEXCOM, INC. | 10/18/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE SENSOR | | Insufficient Information |
| MW5147 086 | 10/12/2023 4:00 | Malfunction | DEXCOM, INC. | 10/18/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE SENSOR | | Insufficient Information |
| MW5147 087 | 10/12/2023 4:00 | Malfunction | DEXCOM, INC. | 10/18/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE SENSOR | Defective Device | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|-----------------|--------------|---|----------------|--|
| MW5146 930 | 10/8/2023 4:00 | Injury | DEXCOM, INC. | 10/12/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia; Hypoglycemia; Diabetic Ketoacidosis |
| MW5146 931 | | Injury | DEXCOM, INC. | 10/12/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5146 871 | | Injury | DEXCOM, INC. | 10/11/2023 4:00 | QBJ | DEXCOM G6 | | Hypoglycemia |
| MW5146 505 | | Injury | DEXCOM, INC. | 10/3/2023 4:00 | QBJ | DEXCOM G6 | | Hypoglycemia |
| MW5146 340 | | Injury | DEXCOM, INC. | 9/29/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5146 355 | | Injury | DEXCOM, INC. | 9/29/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5146 362 | | Injury | DEXCOM, INC. | 9/29/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5146 370 | | Injury | DEXCOM, INC. | 9/29/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5146 223 | 9/25/2023 4:00 | Malfunction | DEXCOM, INC. | 9/27/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5146 224 | 9/25/2023 4:00 | Malfunction | DEXCOM, INC. | 9/27/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|----------------|--------------|--|----------------|---|
| MW5146 225 | 9/25/2023 4:00 | Malfunction | DEXCOM, INC. | 9/27/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR TRANSMITTER | | No Clinical Signs, Symptoms or Conditions |
| MW5146 203 | | Injury | DEXCOM, INC. | 9/26/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5146 144 | 9/20/2023 4:00 | Injury | DEXCOM, INC. | 9/25/2023 4:00 | QBJ | DEXCOM G6 | | Insufficient Information |
| MW5146 150 | | Injury | DEXCOM, INC. | 9/25/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5146 110 | | Injury | DEXCOM, INC. | 9/22/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5145 980 | | Injury | DEXCOM, INC. | 9/20/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5145 922 | | Injury | DEXCOM, INC. | 9/19/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5145 630 | 9/9/2023 4:00 | Injury | DEXCOM, INC. | 9/13/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5145 634 | 9/9/2023 4:00 | Injury | DEXCOM, INC. | 9/13/2023 4:00 | QBJ | DEXCOM | | Skin Inflammation/Irritation |
| MW5145 381 | | Injury | DEXCOM, INC. | 9/7/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5145 386 | | Injury | DEXCOM, INC. | 9/7/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5145 275 | | Injury | DEXCOM, INC. | 9/6/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|------------|--------------|----------------|--------------|------------------|-------------------------------|--|
| MW5144 951 | | Injury | DEXCOM, INC. | 8/24/2023 4:00 | QBJ | DEXCOM G6 | | Hypoglycemia |
| MW5144 952 | | Injury | DEXCOM, INC. | 8/24/2023 4:00 | QBJ | DEXCOM G6 | | Hypoglycemia |
| MW5144 652 | | Injury | DEXCOM, INC. | 8/15/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5144 656 | | Injury | DEXCOM, INC. | 8/15/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5144 657 | | Injury | DEXCOM, INC. | 8/15/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5122 881 | 7/25/2023 4:00 | Injury | DEXCOM, INC. | 8/9/2023 4:00 | QBJ | DEXCOM G6 | | Dehydration; Diarrhea; Fatigue; Hyperglycemia; Vomiting; Diabetic Ketoacidosis |
| MW5120 629 | | Injury | DEXCOM, INC. | 8/4/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5120 350 | | Injury | DEXCOM, INC. | 7/28/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5120 217 | | Injury | DEXCOM, INC. | 7/26/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5119 983 | | Injury | DEXCOM, INC. | 7/21/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5119 229 | 7/5/2023 4:00 | Injury | DEXCOM, INC. | 7/7/2023 4:00 | QBJ | DEXCOM G6 SENSOR | Pumping Stopped; Low Readings | Hyperglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|----------------|--------------|------------------|---|--|
| MW5118 823 | 6/21/2023 4:00 | Injury | DEXCOM, INC. | 6/23/2023 4:00 | MDS | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Therapeutic or Diagnostic Output Failure | Hypoglycemia |
| MW5118 379 | | Injury | DEXCOM, INC. | 6/13/2023 4:00 | QBJ | DEXCOM G6 | Low Readings; Calibration Problem | No Clinical Signs, Symptoms or Conditions |
| MW5118 353 | | Injury | DEXCOM, INC. | 6/12/2023 4:00 | QBJ | DEXCOM G6 | Low Readings | Hyperglycemia |
| MW5118 227 | 6/2/2023 4:00 | Injury | DEXCOM, INC. | 6/6/2023 4:00 | MDS | DEXCOM G6 SENSOR | High Readings | No Clinical Signs, Symptoms or Conditions |
| MW5118 234 | 6/4/2023 4:00 | Malfunction | DEXCOM, INC. | 6/6/2023 4:00 | QBJ | DEXCOM G6 | Loss of or Failure to Bond | No Clinical Signs, Symptoms or Conditions |
| MW5118 248 | 6/3/2023 4:00 | Injury | DEXCOM, INC. | 6/6/2023 4:00 | QBJ | DEXCOM G6 SENSOR | Patient Device Interaction Problem | Erythema; Itching Sensation; Pain; Rash; Fluid Discharge; Skin Inflammation/Irritation; Swelling/Edema |
| MW5118 087 | | Injury | DEXCOM, INC. | 6/1/2023 4:00 | MDS | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |
| MW5118 102 | | Injury | DEXCOM, INC. | 6/1/2023 4:00 | QBJ | DEXCOM G6 | Chemical Problem; Patient Device Interaction Problem | Caustic/Chemical Burns |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|----------------|--------------|------------------|--|---|
| MW5118 033 | 5/28/2023 4:00 | Injury | DEXCOM, INC. | 5/31/2023 4:00 | QBJ | DEXCOM G6 | High Readings | Malaise |
| MW5117 987 | | Injury | DEXCOM, INC. | 5/30/2023 4:00 | QBJ | DEXCOM G6/G7 | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5117 994 | | Injury | DEXCOM INC. | 5/30/2023 4:00 | MDS | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Failure to Analyze Signal; Wireless Communication Problem | Hypoglycemia |
| MW5117 996 | 5/16/2023 4:00 | Injury | DEXCOM, INC. | 5/30/2023 4:00 | QBJ | DEXCOM G6 | Failure to Calibrate | Hyperglycemia; Hypoglycemia |
| MW5117 969 | 5/25/2023 4:00 | Injury | DEXCOM, INC. | 5/26/2023 4:00 | MDS | DEXCOM G6 SENSOR | High Readings | Hypoglycemia |
| MW5117 806 | 5/22/2023 4:00 | Injury | DEXCOM, INC. | 5/24/2023 4:00 | MDS | DEXCOM G6 SENSOR | Product Quality Problem | Insufficient Information |
| MW5117 807 | 5/22/2023 4:00 | Malfunction | DEXCOM, INC. | 5/24/2023 4:00 | MDS | DEXCOM G6 SENSOR | Defective Device | No Clinical Signs, Symptoms or Conditions |
| MW5117 795 | 5/20/2023 4:00 | Injury | DEXCOM, INC. | 5/23/2023 4:00 | MDS | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |
| MW5117 771 | | Malfunction | DEXCOM, INC. | 5/22/2023 4:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|----------------|--------------|--------------------------------------|---|---|
| MW5117 730 | 5/17/2023 4:00 | Malfunction | DEXCOM INC. | 5/19/2023 4:00 | MDS | DEXCOM G6 SENSOR | Low Readings | No Clinical Signs, Symptoms or Conditions |
| MW5117 291 | 4/30/2023 4:00 | Injury | DEXCOM, INC. | 5/5/2023 4:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5117 292 | 4/30/2023 4:00 | Injury | DEXCOM, INC. | 5/5/2023 4:00 | MDS | DEXCOM G6 SENSOR | Entrapment of Device; Incorrect, Inadequate or Imprecise Result or Readings; Connection Problem | Hyperglycemia |
| MW5117 293 | 4/30/2023 4:00 | Injury | DEXCOM, INC. | 5/5/2023 4:00 | MDS | DEXCOM G6 SENSOR | Loss of or Failure to Bond; Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5117 095 | 4/21/2023 4:00 | Injury | DEXCOM, INC. | 4/27/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Device Displays Incorrect Message | No Clinical Signs, Symptoms or Conditions |
| MW5117 097 | 4/21/2021 4:00 | Injury | DEXCOM, INC. | 4/27/2023 4:00 | QBJ | DEXCOM G6 | Device Sensing Problem | Hyperglycemia; Hypoglycemia |
| MW5115 443 | 2/28/2023 5:00 | Malfunction | DEXCOM, INC. | 3/2/2023 5:00 | QBJ | DEXCOM G6 | Application Program Problem; Loss of Data | No Clinical Signs, Symptoms or Conditions |
| MW5114 604 | 1/24/2023 5:00 | Injury | DEXCOM, INC. | 1/31/2023 5:00 | QBJ | DEXCOM G6 APP | Defective Device | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-------------------|------------|-----------------|--------------------|--------------|---|--|--|
| MW5114 583 | 1/27/2023 5:00 | Injury | DEXCOM, INC. | 1/30/2023 5:00 | QBJ | DEXCOM G6 | Delayed Alarm | Hypoglycemia |
| MW5113 775 | 12/9/2022 5:00 | Injury | DEXCOM, INC. | 12/13/2022 5:00 | MDS | DEXCOM G6 | Patient-Device Incompatibility | Erythema; Hemorrhage/Bleedi ng; Itching Sensation; Rash; Skin Inflammation/ Irritation |
| MW5112 549 | 10/6/2022 4:00 | Injury | DEXCOM, INC. | 10/11/2022 4:00 | QBJ | DEXCOM G6 CONTINUO US GLUCOSE METER | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5112 549 | 10/6/2022 4:00 | Injury | DEXCOM, INC. | 10/11/2022 4:00 | QBJ | SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5112 302 | 9/10/2022 4:00 | Injury | DEXCOM, INC. | 9/27/2022 4:00 | QBJ | DEXCOM G6 | High Readings | Crushing Injury; Hypoglycemia |
| MW5112 031 | 6/15/2022 4:00 | Injury | DEXCOM, INC. | 9/13/2022 4:00 | QBJ | DEXCOM G6 TRANSMITT ER | Loss of Data | Insufficient Information |
| MW5111 831 | 8/24/2022 4:00 | Injury | DEXCOM, INC. | 8/31/2022 4:00 | QBJ | DEXCOM G6 SENSOR AND TRANSMITT ER | Output Problem | Insufficient Information |
| MW5111 683 | 8/20/2022 4:00 | Injury | DEXCOM, INC. | 8/23/2022 4:00 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Hypersensitivity/All ergic reaction; Unspecified Infection; Rash; Skin Inflammation/ Irritation |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-------------------|-------------|--------------|-------------------|--------------|-----------------------|--|--|
| MW5111422 | 8/5/2022 4:00 | Injury | DEXCOM, INC. | 8/10/2022 4:00 | QBJ | DEXCOM G6 APPLICATION | Product Quality Problem; Application Program Version or Upgrade Problem | Hypoglycemia |
| MW5111054 | 7/16/2022 4:00 | Injury | DEXCOM, INC. | 7/21/2022 4:00 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Erythema; Itching Sensation; Rash |
| MW5110452 | 6/17/2022 4:00 | Malfunction | DEXCOM, INC. | 6/22/2022 4:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Calibration Problem | Insufficient Information |
| MW5110333 | 6/11/2022 4:00 | Injury | DEXCOM, INC. | 6/14/2022 4:00 | MDS | DEXCOM G6 SENSORS | Therapeutic or Diagnostic Output Failure | Hypoglycemia |
| MW5110333 | 6/11/2022 4:00 | Injury | DEXCOM, INC. | 6/14/2022 4:00 | MDS | DEXCOM G6 SENSORS | Therapeutic or Diagnostic Output Failure | Hypoglycemia |
| MW5110333 | 6/11/2022 4:00 | Injury | DEXCOM, INC. | 6/14/2022 4:00 | MDS | DEXCOM G6 SENSORS | Therapeutic or Diagnostic Output Failure | Hypoglycemia |
| MW5110333 | 6/11/2022 4:00 | Injury | DEXCOM, INC. | 6/14/2022 4:00 | MDS | DEXCOM G6 SENSORS | Therapeutic or Diagnostic Output Failure | Hypoglycemia |
| MW5109768 | 4/6/2022 4:00 | Malfunction | DEXCOM, INC. | 5/17/2022 4:00 | QBJ | DEXCOM G6 SENSOR | Difficult or Delayed Activation | Rash |
| MW5109624 | 4/1/2022 4:00 | Injury | DEXCOM, INC. | 5/10/2022 4:00 | QBJ | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Erythema; Peeling; Blister; Skin Inflammation/Irritation |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|----------------|--------------|---|--|---|
| MW5109 507 | 4/10/2022 4:00 | Injury | DEXCOM, INC. | 5/3/2022 4:00 | QBJ | DEXCOM G6 CGM | Patient-Device Incompatibility | Erythema; Itching Sensation; Local Reaction |
| MW5109 507 | 4/10/2022 4:00 | Injury | DEXCOM, INC. | 5/3/2022 4:00 | QBJ | DEXCOM G6 CGM | Patient-Device Incompatibility | Erythema; Itching Sensation; Local Reaction |
| MW5109 166 | 4/15/2022 4:00 | Injury | DEXCOM, INC. | 4/19/2022 4:00 | QDK | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Erythema; Rash; Skin Tears |
| MW5109 180 | 4/2/2022 4:00 | Malfunction | DEXCOM, INC. | 4/19/2022 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Device Markings/Labelling Problem | Hyperglycemia |
| MW5108 611 | | Malfunction | DEXCOM, INC. | 3/29/2022 4:00 | MDS | DEXCOM G6 | Low Test Results; Material Integrity Problem | Insufficient Information |
| MW5107 031 | | Injury | DEXCOM, INC. | 1/26/2022 5:00 | QBJ | DEXCOM CONTINUOUS GLUCOSE MONITORING SYSTEM TRANSMITTER | Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem | Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation |
| MW5107 031 | | Injury | DEXCOM, INC. | 1/26/2022 5:00 | QBJ | DEXCOM CONTINUOUS GLUCOSE MONITORING SYSTEM TRANSMITTER | Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem | Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|--------------|----------------|--------------|---|--|---|
| MW5107031 | | Injury | DEXCOM, INC. | 1/26/2022 5:00 | QBJ | DEXCOM CONTINUOUS GLUCOSE MONITORING SYSTEM TRANSMITTER | Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem | Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation |
| MW5107031 | | Injury | DEXCOM, INC. | 1/26/2022 5:00 | QBJ | SENSORS | Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem | Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation |
| MW5107031 | | Injury | DEXCOM, INC. | 1/26/2022 5:00 | QBJ | SENSORS | Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem | Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation |
| MW5107031 | | Injury | DEXCOM, INC. | 1/26/2022 5:00 | QBJ | SENSORS | Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem | Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation |
| MW5107031 | | Injury | DEXCOM, INC. | 1/26/2022 5:00 | QBJ | SENSORS | Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem | Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|----------------|--------------|--------------------------------------|--|---|
| MW5107031 | | Injury | DEXCOM, INC. | 1/26/2022 5:00 | QBJ | SENSORS | Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem | Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation |
| MW5106846 | 1/17/2022 5:00 | Injury | DEXCOM, INC. | 1/20/2022 5:00 | QBJ | DEXCOM G6 | Product Quality Problem; Patient Device Interaction Problem | No Clinical Signs, Symptoms or Conditions |
| MW5106846 | 1/17/2022 5:00 | Injury | DEXCOM, INC. | 1/20/2022 5:00 | QBJ | SENSOR | Product Quality Problem; Patient Device Interaction Problem | No Clinical Signs, Symptoms or Conditions |
| MW5106533 | 1/1/2022 5:00 | Malfunction | DEXCOM, INC. | 1/7/2022 5:00 | QBJ | DEXCOM G6 TRANSMITTERS | Defective Device; Device Markings/Labelling Problem | Insufficient Information |
| MW5106533 | 1/1/2022 5:00 | Malfunction | DEXCOM, INC. | 1/7/2022 5:00 | QBJ | DEXCOM G6 TRANSMITTERS | Defective Device; Device Markings/Labelling Problem | Insufficient Information |
| MW5106534 | 1/5/2022 5:00 | Malfunction | DEXCOM, INC. | 1/7/2022 5:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Product Quality Problem; Incorrect, Inadequate or Imprecise Result or Readings; Output Problem | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-----------------|------------|--------------|-----------------|--------------|---------------|---|------------------------------|
| MW5106 453 | 12/30/2021 5:00 | Injury | DEXCOM, INC. | 1/4/2022 5:00 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Rash; Caustic/Chemical Burns |
| MW5106 223 | 12/18/2021 5:00 | Injury | DEXCOM, INC. | 12/21/2021 5:00 | MDS | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Hyperglycemia |
| MW5106 223 | 12/18/2021 5:00 | Injury | DEXCOM, INC. | 12/21/2021 5:00 | MDS | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Hyperglycemia |
| MW5106 223 | 12/18/2021 5:00 | Injury | DEXCOM, INC. | 12/21/2021 5:00 | MDS | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Hyperglycemia |
| MW5106 223 | 12/18/2021 5:00 | Injury | DEXCOM, INC. | 12/21/2021 5:00 | MDS | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Hyperglycemia |
| MW5106 223 | 12/18/2021 5:00 | Injury | DEXCOM, INC. | 12/21/2021 5:00 | MDS | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Hyperglycemia |
| MW5106 187 | 12/16/2021 5:00 | Injury | DEXCOM, INC. | 12/20/2021 5:00 | QBJ | DEXCOM G6 CGM | Device Markings/Labelling Problem | Insufficient Information |
| MW5106 187 | 12/16/2021 5:00 | Injury | DEXCOM, INC. | 12/20/2021 5:00 | QBJ | SENSOR | Device Markings/Labelling Problem | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-----------------|-------------|--------------|-----------------|--------------|--|--|---|
| MW5106 133 | 11/1/2020 4:00 | Injury | DEXCOM INC | 12/16/2021 5:00 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Erythema; Itching Sensation; Rash; Blister |
| MW5106 145 | 12/15/2021 5:00 | Malfunction | DEXCOM, INC | 12/16/2021 5:00 | QBJ | DEXCOM G6 CGM 10 DAY ON BODY CONTINUOUS GLUCOSE SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5106 145 | 12/15/2021 5:00 | Malfunction | DEXCOM, INC | 12/16/2021 5:00 | QBJ | TRANSMITTER | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5105 309 | 11/9/2021 5:00 | Injury | DEXCOM, INC. | 11/12/2021 5:00 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Rash |
| MW5105 309 | 11/9/2021 5:00 | Injury | DEXCOM, INC. | 11/12/2021 5:00 | QBJ | SENSOR | Patient-Device Incompatibility | Rash |
| MW5105 241 | 11/2/2021 4:00 | Injury | DEXCOM, INC. | 11/9/2021 5:00 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Bruise/Contusion; Purulent Discharge; Hemorrhage/Bleeding; Itching Sensation; Peeling; Skin Discoloration; Skin Inflammation/Irritation |
| MW5105 247 | 11/6/2021 4:00 | Injury | DEXCOM, INC. | 11/9/2021 5:00 | QBJ | DEXCOM G6 | Patient Device Interaction Problem | Hypoglycemia |
| MW5105 095 | 10/1/2021 4:00 | Injury | DEXCOM, INC. | 11/2/2021 4:00 | QBJ | DEXCOM G6 APP FOR ANDROID | Labelling, Instructions for Use or Training Problem; Product Quality Problem | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|--|--|---|
| MW5104 943 | 10/15/2021 | Injury | DEXCOM, INC. | 10/25/2021 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITORING SYSTEM | Defective Alarm; Computer Software Problem; Device Sensing Problem | Hyperglycemia |
| MW5104 775 | 10/15/2021 | Malfunction | DEXCOM, INC. | 10/19/2021 | MDS | DEXCOM G6 GLUCOSE SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5104 775 | 10/15/2021 | Malfunction | DEXCOM, INC. | 10/19/2021 | MDS | DEXCOM G6 GLUCOSE SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5104 775 | 10/15/2021 | Malfunction | DEXCOM, INC. | 10/19/2021 | MDS | DEXCOM G6 GLUCOSE SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5104 775 | 10/15/2021 | Malfunction | DEXCOM, INC. | 10/19/2021 | QBJ | TRANSMITTER | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5104 732 | 1/1/2021 | Injury | DEXCOM, INC. | 10/18/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | Patient-Device Incompatibility; Therapeutic or Diagnostic Output Failure | Scar Tissue; Skin Disorders; Skin Inflammation/Irritation |
| MW5104 311 | 9/21/2021 | Malfunction | DEXCOM INC. | 9/30/2021 | QBJ | DEXCOM G6 CGM | Communication or Transmission Problem; Device-Device Incompatibility; Therapeutic or Diagnostic Output Failure | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|-----------------------|---|--|
| MW5104 208 | 9/23/2021 | Injury | DEXCOM, INC. | 9/24/2021 | MDS | DEXCOM G6 SENSOR | Defective Component | No Clinical Signs, Symptoms or Conditions |
| MW5104 208 | 9/23/2021 | Injury | DEXCOM, INC. | 9/24/2021 | QBJ | TRANSMITTER | Defective Component | No Clinical Signs, Symptoms or Conditions |
| MW5103 803 | 8/31/2021 | Injury | DEXCOM, INC. | 9/8/2021 | MDS | DEXCOM G6 SENSOR | Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5103 803 | 8/31/2021 | Injury | DEXCOM, INC. | 9/8/2021 | MDS | DEXCOM G6 SENSOR | Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5103 803 | 8/31/2021 | Injury | DEXCOM, INC. | 9/8/2021 | MDS | DEXCOM G6 SENSOR | Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5103 716 | 8/31/2021 | Injury | DEXCOM, INC. | 9/2/2021 | MDS | DEXCOM G6 | Defective Component | Emotional Changes |
| MW5103 674 | 8/19/2021 | Injury | DEXCOM, INC. | 9/1/2021 | QBJ | DEXCOM G6 | Product Quality Problem; Incorrect, Inadequate or Imprecise Result or Readings; Defective Device; Calibration Problem | Apnea; Cardiac Arrest; Loss of consciousness; Hypoglycemic Shock |
| MW5103 457 | | Malfunction | DEXCOM G6 | 8/23/2021 | QBJ | DEXCOM G6 TRANSMITTER | Use of Device Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|-----------------------|---|--------------------------|
| MW5103 457 | | Malfunction | DEXCOM G6 | 8/23/2021 | QBJ | DEXCOM G6 TRANSMITTER | Use of Device Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5103 457 | | Malfunction | DEXCOM G6 | 8/23/2021 | QBJ | DEXCOM G6 TRANSMITTER | Use of Device Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5103 457 | | Malfunction | DEXCOM G6 | 8/23/2021 | QBJ | DEXCOM G6 TRANSMITTER | Use of Device Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5103 457 | | Malfunction | DEXCOM G6 | 8/23/2021 | QBJ | DEXCOM G6 TRANSMITTER | Use of Device Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5103 457 | | Malfunction | DEXCOM G6 | 8/23/2021 | QBJ | DEXCOM G6 TRANSMITTER | Use of Device Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5103 457 | | Malfunction | DEXCOM G6 | 8/23/2021 | QBJ | DEXCOM G6 TRANSMITTER | Use of Device Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5103 457 | | Malfunction | DEXCOM G6 | 8/23/2021 | QBJ | DEXCOM G6 TRANSMITTER | Use of Device Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5103 185 | 5/19/2021 | Injury | DEXCOM, INC. | 8/11/2021 | MDS | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | Hypoglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|-------------------------------------|--|--------------------------|
| MW5103054 | 8/4/2021 | Malfunction | DEXCOM, INC. | 8/5/2021 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |
| MW5103054 | 8/4/2021 | Malfunction | DEXCOM, INC. | 8/5/2021 | MDS | SENSORS | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |
| MW5102993 | 8/1/2021 | Injury | DEXCOM, INC. | 8/3/2021 | MDS | DEXCOM G6 CONTINUAL GLUCOSE MONITOR | False Alarm; Low Readings; Therapeutic or Diagnostic Output Failure | Hyperglycemia |
| MW5102502 | 7/8/2021 | Injury | DEXCOM, INC. | 7/12/2021 | MDS | DEXCOM G6 SENSOR | Defective Component; Separation Failure; Physical Resistance/Sticking | Insufficient Information |
| MW5102052 | 6/19/2021 | Malfunction | DEXCOM, INC. | 6/22/2021 | QBJ | DEXCOM G6 TRANSMITTER | Expiration Date Error; Device Markings/Labelling Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5102052 | 6/19/2021 | Malfunction | DEXCOM, INC. | 6/22/2021 | QBJ | DEXCOM G6 TRANSMITTER | Expiration Date Error; Device Markings/Labelling Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|-----------------------|--|---|
| MW5102052 | 6/19/2021 | Malfunction | DEXCOM, INC. | 6/22/2021 | QBJ | DEXCOM G6 TRANSMITTER | Expiration Date Error; Device Markings/Labelling Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5102052 | 6/19/2021 | Malfunction | DEXCOM, INC. | 6/22/2021 | MDS | SENSORS | Expiration Date Error; Device Markings/Labelling Problem; Therapeutic or Diagnostic Output Failure | Insufficient Information |
| MW5102054 | 6/22/2021 | Injury | DEXCOM, INC. | 6/22/2021 | QBJ | DEXCOM CGM | Retraction Problem | Insufficient Information |
| MW5102054 | 6/22/2021 | Injury | DEXCOM, INC. | 6/22/2021 | QBJ | DEXCOM CGM | Retraction Problem | Insufficient Information |
| MW5101707 | | Malfunction | DEXCOM, INC. | 6/3/2021 | QBJ | DEXCOM G6 | Defective Device | No Clinical Signs, Symptoms or Conditions |
| MW5101707 | | Malfunction | DEXCOM, INC. | 6/3/2021 | MDS | SENSOR | Defective Device | No Clinical Signs, Symptoms or Conditions |
| MW5101606 | 5/25/2021 | Injury | DEXCOM, INC. | 5/27/2021 | QBJ | DEXCOM G6 | Device Ingredient or Reagent Problem | Partial thickness (Second Degree) Burn |
| MW5101606 | 5/25/2021 | Injury | DEXCOM, INC. | 5/27/2021 | MDS | DEXCOM G6 | Device Ingredient or Reagent Problem | Partial thickness (Second Degree) Burn |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|---------------------------|---|--------------------------|
| MW5101 373 | 5/12/2021 | Malfunction | DEXCOM INC. | 5/14/2021 | QBJ | DEXCOM G6 SENSOR INSERTER | Activation, Positioning or Separation Problem | Insufficient Information |
| MW5101 154 | 4/30/2021 | Malfunction | DEXCOM, INC. | 5/4/2021 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Insufficient Information |
| MW5101 154 | 4/30/2021 | Malfunction | DEXCOM, INC. | 5/4/2021 | MDS | SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Insufficient Information |
| MW5101 154 | 4/30/2021 | Malfunction | DEXCOM, INC. | 5/4/2021 | MDS | SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Insufficient Information |
| MW5101 154 | 4/30/2021 | Malfunction | DEXCOM, INC. | 5/4/2021 | MDS | SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Insufficient Information |
| MW5101 154 | 4/30/2021 | Malfunction | DEXCOM, INC. | 5/4/2021 | MDS | SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|-----------------------|---|---|
| MW5101154 | 4/30/2021 | Malfunction | DEXCOM, INC. | 5/4/2021 | MDS | SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Insufficient Information |
| MW5101103 | 4/28/2021 | Malfunction | DEXCOM, INC. | 4/30/2021 | QBJ | DEXCOM G6 TRANSMITTER | Incorrect, Inadequate or Imprecise Result or Readings; Defective Device | Hypoglycemia |
| MW5101103 | 4/28/2021 | Malfunction | DEXCOM, INC. | 4/30/2021 | QBJ | DEXCOM G6 TRANSMITTER | Incorrect, Inadequate or Imprecise Result or Readings; Defective Device | Hypoglycemia |
| MW5101021 | 8/24/2020 | Injury | DEXCOM, INC. | 4/27/2021 | QBJ | DEXCOM G6 CGM | Detachment of Device or Device Component | Foreign Body In Patient |
| MW5101021 | 8/24/2020 | Injury | DEXCOM, INC. | 4/27/2021 | MDS | SENSOR | Detachment of Device or Device Component | Foreign Body In Patient |
| MW5101049 | 4/1/2020 | Injury | DEXCOM, INC. | 4/27/2021 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Hypersensitivity/Allergic reaction; Unspecified Infection |
| MW5100898 | 4/11/2021 | Injury | DEXCOM, INC. | 4/20/2021 | QBJ | DEXCOM G6 | Use of Device Problem; Battery Problem; Complete Loss of Power | Hyperglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|----------------------|--|--|
| MW5100 853 | 4/14/2021 | Malfunction | DEXCOM, INC. | 4/19/2021 | MDS | DEXCOM G6 SENSOR | Defective Device; Data Problem | Insufficient Information |
| MW5100 853 | 4/14/2021 | Malfunction | DEXCOM, INC. | 4/19/2021 | MDS | DEXCOM G6 SENSOR | Defective Device; Data Problem | Insufficient Information |
| MW5100 853 | 4/14/2021 | Malfunction | DEXCOM, INC. | 4/19/2021 | MDS | DEXCOM G6 SENSOR | Defective Device; Data Problem | Insufficient Information |
| MW5100 853 | 4/14/2021 | Malfunction | DEXCOM, INC. | 4/19/2021 | MDS | DEXCOM G6 SENSOR | Defective Device; Data Problem | Insufficient Information |
| MW5100 696 | 4/7/2021 | Injury | DEXCOM, INC | 4/9/2021 | MDS | DEXCOM G6 CGM SENSOR | Patient Device Interaction Problem | Hypersensitivity/Allergic reaction; Rash |
| MW5100 649 | 4/2/2021 | Injury | DEXCOM INC | 4/7/2021 | QBJ | DEXCOM G6 | Use of Device Problem; Improper or Incorrect Procedure or Method; Separation Failure | Insufficient Information |
| MW5100 046 | 1/1/2021 | Injury | DEXCOM, INC. | 3/16/2021 | QBJ | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Hypersensitivity/Allergic reaction; Caustic/Chemical Burns |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|--------------|---------------|--------------|------------------|--|--|
| MW5100007 | 2/19/2021 | Injury | DEXCOM, INC. | 3/15/2021 | MDS | DEXCOM G6 SENSOR | Loss of or Failure to Bond; Incorrect, Inadequate or Imprecise Result or Readings; Patient-Device Incompatibility; Device Fell | Itching Sensation; Pain; Rash; Scar Tissue; Skin Inflammation/Irritation |
| MW5100007 | 2/19/2021 | Injury | DEXCOM, INC. | 3/15/2021 | MDS | DEXCOM G6 SENSOR | Loss of or Failure to Bond; Incorrect, Inadequate or Imprecise Result or Readings; Patient-Device Incompatibility; Device Fell | Itching Sensation; Pain; Rash; Scar Tissue; Skin Inflammation/Irritation |
| MW5100007 | 2/19/2021 | Injury | DEXCOM, INC. | 3/15/2021 | MDS | DEXCOM G6 SENSOR | Loss of or Failure to Bond; Incorrect, Inadequate or Imprecise Result or Readings; Patient-Device Incompatibility; Device Fell | Itching Sensation; Pain; Rash; Scar Tissue; Skin Inflammation/Irritation |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|--------------|---------------|--------------|------------------|--|--|
| MW5100007 | 2/19/2021 | Injury | DEXCOM, INC. | 3/15/2021 | MDS | DEXCOM G6 SENSOR | Loss of or Failure to Bond; Incorrect, Inadequate or Imprecise Result or Readings; Patient-Device Incompatibility; Device Fell | Itching Sensation; Pain; Rash; Scar Tissue; Skin Inflammation/Irritation |
| MW5100008 | 3/1/2021 | Injury | DEXCOM, INC. | 3/15/2021 | MDS | DEXCOM G6 | Patient-Device Incompatibility | Erythema; Itching Sensation; Superficial (First Degree) Burn |
| MW5099926 | 3/8/2021 | Injury | DEXCOM, INC. | 3/10/2021 | QBJ | DEXCOM G6 CGM | Entrapment of Device; Separation Failure; Physical Resistance/Sticking | Pain |
| MW5099926 | 3/8/2021 | Injury | DEXCOM, INC. | 3/10/2021 | MDS | SENSOR | Entrapment of Device; Separation Failure; Physical Resistance/Sticking | Pain |
| MW5099769 | 3/1/2021 | Injury | DEXCOM, INC. | 3/3/2021 | MDS | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Rash |
| MW5099774 | 2/15/2021 | Injury | DEXCOM, INC | 3/3/2021 | QBJ | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Erythema; Itching Sensation; Swelling |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|------------|--------------|----------------|--------------|---|--|---|
| MW5099 774 | 2/15/2021 | Injury | DEXCOM, INC. | 3/3/2021 | MDS | SENSOR | Adverse Event Without Identified Device or Use Problem | Erythema; Itching Sensation; Swelling |
| MW5099 749 | 2/24/2021 | Injury | DEXCOM INC. | 3/2/2021 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Reaction |
| MW5099 721 | 2/25/2021 | Injury | DEXCOM, INC. | 3/1/2021 | MDS | DEXCOM G6 CGM | Adverse Event Without Identified Device or Use Problem | Itching Sensation; Rash; Burning Sensation; Blister |
| MW5099 627 | 2/1/2021 | Injury | DEXCOM INC. | 2/24/2021 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Itching Sensation; Rash; Skin Irritation; Swelling |
| MW5099 627 | 2/1/2021 | Injury | DEXCOM INC. | 2/24/2021 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Itching Sensation; Rash; Skin Irritation; Swelling |
| MW5099 627 | 2/1/2021 | Injury | DEXCOM INC. | 2/24/2021 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Itching Sensation; Rash; Skin Irritation; Swelling |
| MW5099 627 | 2/1/2021 | Injury | DEXCOM INC. | 2/24/2021 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Itching Sensation; Rash; Skin Irritation; Swelling |
| MW5099 574 | 11/2/2022 4:00 | Injury | DEXCOM, INC. | 2/23/2021 5:00 | QBJ | DEXCOM G6 CONTINUOUS US GLUCOSE MONITOR | Inaccurate Delivery; Unexpected Shutdown | Hypoglycemia; Overdose; Diabetic Ketoacidosis |
| MW5099 574 | 11/2/2022 4:00 | Injury | DEXCOM, INC. | 2/23/2021 5:00 | QBJ | DEXCOM G6 CONTINUOUS US GLUCOSE MONITOR | Inaccurate Delivery; Unexpected Shutdown | Hypoglycemia; Overdose; Diabetic Ketoacidosis |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|------------|--------------|----------------|--------------|------------------------------------|--|---|
| MW5099 574 | 11/2/2022 4:00 | Injury | DEXCOM, INC. | 2/23/2021 5:00 | QBJ | DEXCOM SENSOR | Inaccurate Delivery; Unexpected Shutdown | Hypoglycemia; Overdose; Diabetic Ketoacidosis |
| MW5099 574 | 11/2/2022 4:00 | Injury | DEXCOM, INC. | 2/23/2021 5:00 | QBJ | DEXCOM SENSOR | Inaccurate Delivery; Unexpected Shutdown | Hypoglycemia; Overdose; Diabetic Ketoacidosis |
| MW5099 578 | 2/21/2021 | Injury | DEXCOM, INC. | 2/23/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE METER | Device Displays Incorrect Message; Communication or Transmission Problem | Hypoglycemia |
| MW5099 578 | 2/21/2021 | Injury | DEXCOM, INC. | 2/23/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE METER | Device Displays Incorrect Message; Communication or Transmission Problem | Hypoglycemia |
| MW5099 578 | 2/21/2021 | Injury | DEXCOM, INC. | 2/23/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE METER | Device Displays Incorrect Message; Communication or Transmission Problem | Hypoglycemia |
| MW5099 578 | 2/21/2021 | Injury | DEXCOM, INC. | 2/23/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE METER | Device Displays Incorrect Message; Communication or Transmission Problem | Hypoglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|------------------------------------|--|---|
| MW5099 578 | 2/21/2021 | Injury | DEXCOM, INC. | 2/23/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE METER | Device Displays Incorrect Message; Communication or Transmission Problem | Hypoglycemia |
| MW5099 578 | 2/21/2021 | Injury | DEXCOM, INC. | 2/23/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE METER | Device Displays Incorrect Message; Communication or Transmission Problem | Hypoglycemia |
| MW5099 578 | 2/21/2021 | Injury | DEXCOM, INC. | 2/23/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE METER | Device Displays Incorrect Message; Communication or Transmission Problem | Hypoglycemia |
| MW5099 578 | 2/21/2021 | Injury | DEXCOM, INC. | 2/23/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE METER | Device Displays Incorrect Message; Communication or Transmission Problem | Hypoglycemia |
| MW5099 578 | 2/21/2021 | Injury | DEXCOM, INC. | 2/23/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE METER | Device Displays Incorrect Message; Communication or Transmission Problem | Hypoglycemia |
| MW5099 534 | 2/17/2021 | Malfunction | DEXCOM, INC. | 2/22/2021 | QBJ | DEXCOM G6 CGM | Separation Failure | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|---|--|--------------------------------------|
| MW5099 497 | 2/16/2021 | Malfunction | DEXCOM, INC. | 2/18/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITORING SYSTEM SENSOR | Activation Failure | Rash; Local Reaction |
| MW5099 497 | 2/16/2021 | Malfunction | DEXCOM, INC. | 2/18/2021 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITORING SYSTEM SENSOR | Activation Failure | Rash; Local Reaction |
| MW5099 459 | | Injury | DEXCOM, INC. | 2/17/2021 | MDS | DEXCOM G6 | Difficult or Delayed Separation; Erratic Results | Rash |
| MW5099 311 | 2/6/2021 | Injury | DEXCOM, INC. | 2/9/2021 | MDS | DEXCOM G6 CGM | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia; Diabetic Ketoacidosis |
| MW5099 311 | 2/6/2021 | Injury | DEXCOM, INC. | 2/9/2021 | QBJ | DEXCOM G6 CGM | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia; Diabetic Ketoacidosis |
| MW5099 325 | 2/8/2021 | Malfunction | DEXCOM, INC. | 2/9/2021 | MDS | DEXCOM G-6 SENSOR | Adverse Event Without Identified Device or Use Problem | Erythema; Itching Sensation; Rash |
| MW5099 325 | 2/8/2021 | Malfunction | DEXCOM, INC. | 2/9/2021 | MDS | DEXCOM G-6 SENSOR | Adverse Event Without Identified Device or Use Problem | Erythema; Itching Sensation; Rash |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|-------------------|--|--|
| MW5099 325 | 2/8/2021 | Malfunction | DEXCOM, INC. | 2/9/2021 | MDS | DEXCOM G-6 SENSOR | Adverse Event Without Identified Device or Use Problem | Erythema; Itching Sensation; Rash |
| MW5099 210 | 1/31/2021 | Injury | DEXCOM, INC. | 2/3/2021 | MDS | DEXCOM G6 | Difficult to Remove | Blister; Skin Inflammation/ Irritation |
| MW5099 159 | 12/16/2020 | Injury | DEXCOM, INC. | 2/2/2021 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Rash |
| MW5099 159 | 12/16/2020 | Injury | DEXCOM, INC. | 2/2/2021 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Rash |
| MW5099 160 | 10/12/2020 | Injury | DEXCOM, INC. | 2/2/2021 | MDS | DEXCOM G6 | Patient-Device Incompatibility | Rash |
| MW5099 160 | 10/12/2020 | Injury | DEXCOM, INC. | 2/2/2021 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Rash |
| MW5099 165 | 1/15/2021 | Injury | DEXCOM, INC. | 2/2/2021 | MDS | DEXCOM G6 | Patient-Device Incompatibility | Unspecified Infection; Burn, Thermal |
| MW5099 166 | 12/30/2020 | Injury | DEXCOM, INC. | 2/2/2021 | MDS | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Itching Sensation; Rash |
| MW5099 166 | 12/30/2020 | Injury | DEXCOM, INC. | 2/2/2021 | MDS | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Itching Sensation; Rash |
| MW5099 166 | 12/30/2020 | Injury | DEXCOM, INC. | 2/2/2021 | MDS | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Itching Sensation; Rash |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|---------------------------------|--|--|
| MW5099 170 | 7/17/2020 | Injury | DEXCOM, INC. | 2/2/2021 | MDS | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Erythema; Inflammation; Itching Sensation; Skin Irritation |
| MW5099 171 | 1/25/2021 | Injury | DEXCOM, INC. | 2/2/2021 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Rash |
| MW5099 114 | 1/1/2021 | Injury | DEXCOM, INC. | 1/29/2021 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Pain; Rash |
| MW5099 021 | 1/23/2021 | Malfunction | DEXCOM, INC. | 1/26/2021 | QBJ | DEXCOM G6 CONTINUOUS GLUCOMETER | Incorrect, Inadequate or Imprecise Result or Readings | Hypoglycemia |
| MW5098 967 | 11/26/2020 | Injury | DEXCOM INC | 1/22/2021 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Itching Sensation; Rash; Skin Irritation |
| MW5098 967 | 11/26/2020 | Injury | DEXCOM INC | 1/22/2021 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Itching Sensation; Rash; Skin Irritation |
| MW5098 967 | 11/26/2020 | Injury | DEXCOM INC | 1/22/2021 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Itching Sensation; Rash; Skin Irritation |
| MW5098 967 | 11/26/2020 | Injury | DEXCOM INC | 1/22/2021 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Itching Sensation; Rash; Skin Irritation |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|------------------|--|--|
| MW50989 67 | 11/26/2020 | Injury | DEXCOM INC | 1/22/2021 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Itching Sensation; Rash; Skin Irritation |
| MW50989 67 | 11/26/2020 | Injury | DEXCOM INC | 1/22/2021 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Itching Sensation; Rash; Skin Irritation |
| MW50989 15 | 1/15/2021 | Injury | DEXCOM INC | 1/20/2021 | MDS | DEXCOM G6 | Therapeutic or Diagnostic Output Failure | Erythema |
| MW50989 15 | 1/15/2021 | Injury | DEXCOM INC | 1/20/2021 | MDS | DEXCOM G6 | Therapeutic or Diagnostic Output Failure | Erythema |
| MW50989 15 | 1/15/2021 | Injury | DEXCOM INC | 1/20/2021 | MDS | DEXCOM G6 | Therapeutic or Diagnostic Output Failure | Erythema |
| MW50989 16 | 11/29/2020 | Injury | DEXCOM INC. | 1/20/2021 | QBJ | DEXCOM G6 | | |
| MW50988 61 | 1/12/2021 | Malfunction | DEXCOM INC. | 1/15/2021 | QBJ | DEXCOM G6 | Retraction Problem; Physical Resistance/Sticking | Pain |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|---|---|-----------------------------|
| MW50988 73 | 1/13/2021 | Malfunction | DEXCOM INC | 1/15/2021 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem | Insufficient Information |
| MW50988 73 | 1/13/2021 | Malfunction | DEXCOM INC | 1/15/2021 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem | Insufficient Information |
| MW50988 73 | 1/13/2021 | Malfunction | DEXCOM INC | 1/15/2021 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|--------------------------------------|--|---------------------------|
| MW5098873 | 1/13/2021 | Malfunction | DEXCOM INC | 1/15/2021 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem | Insufficient Information |
| MW5098873 | 1/13/2021 | Malfunction | DEXCOM INC | 1/15/2021 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem | Insufficient Information |
| MW5098873 | 1/13/2021 | Malfunction | DEXCOM INC | 1/15/2021 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem | Insufficient Information |
| MW5098827 | 1/4/2021 | Injury | DEXCOM, INC. | 1/13/2021 | MDS | DEXCOM G6 SENSOR | Mechanical Problem | Hemorrhage/Bleeding; Pain |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|--------------|---------------|--------------|--------------------------------------|--|--|
| MW5098719 | 5/1/2020 | Injury | DEXCOM INC | 1/6/2021 | MDS | DEXCOM G6 CGM | Patient-Device Incompatibility | Itching Sensation; Pain; Scarring; Reaction; Caustic/Chemical Burns |
| MW5098608 | 9/9/2020 | Injury | DEXCOM, INC. | 12/31/2020 | MDS | DEXCOM G6 CGM | Patient-Device Incompatibility | Purulent Discharge; Hemorrhage/Bleeding; Itching Sensation; Caustic/Chemical Burns |
| MW5098608 | 9/9/2020 | Injury | DEXCOM, INC. | 12/31/2020 | MDS | DEXCOM G6 CGM | Patient-Device Incompatibility | Purulent Discharge; Hemorrhage/Bleeding; Itching Sensation; Caustic/Chemical Burns |
| MW5098614 | | Injury | DEXCOM, INC. | 12/31/2020 | QBJ | DEXCOM G6 | Application Program Problem; Device-Device Incompatibility | Insufficient Information |
| MW5098567 | 12/25/2020 | Injury | DEXCOM, INC. | 12/29/2020 | MDS | DEXCOM G6 SENSOR | Separation Failure | Hyperglycemia |
| MW5098567 | 12/25/2020 | Injury | DEXCOM, INC. | 12/29/2020 | MDS | DEXCOM G6 SENSOR | Separation Failure | Hyperglycemia |
| MW5098457 | 12/5/2020 | Injury | DEXCOM | 12/18/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Retraction Problem; Separation Failure | Pain |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|------------------|--|---|
| MW5098396 | 12/12/2020 | Malfunction | DEXCOM INC. | 12/15/2020 | QBJ | DEXCOM G6 | No Audible Alarm | Insufficient Information |
| MW5098407 | 5/3/2020 | Injury | DEXCOM INC. | 12/15/2020 | MDS | DEXCOM G6 SENSOR | Detachment of Device or Device Component | Hemorrhage/Bleeding; Itching Sensation; Rash; Blister |
| MW5098407 | 5/3/2020 | Injury | DEXCOM INC. | 12/15/2020 | MDS | DEXCOM G6 SENSOR | Detachment of Device or Device Component | Hemorrhage/Bleeding; Itching Sensation; Rash; Blister |
| MW5098407 | 5/3/2020 | Injury | DEXCOM INC. | 12/15/2020 | MDS | DEXCOM G6 SENSOR | Detachment of Device or Device Component | Hemorrhage/Bleeding; Itching Sensation; Rash; Blister |
| MW5098407 | 5/3/2020 | Injury | DEXCOM INC. | 12/15/2020 | MDS | DEXCOM G6 SENSOR | Detachment of Device or Device Component | Hemorrhage/Bleeding; Itching Sensation; Rash; Blister |
| MW5098407 | 5/3/2020 | Injury | DEXCOM INC. | 12/15/2020 | MDS | DEXCOM G6 SENSOR | Detachment of Device or Device Component | Hemorrhage/Bleeding; Itching Sensation; Rash; Blister |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|-----------------|---------------|--------------|---|---|--|
| MW50983 30 | 11/22/2020 | Injury | DEXCOM, INC. | 12/11/2020 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITORING (CGM) SYSTEM | Patient-Device Incompatibility | Local Reaction; Scarring; Burning Sensation |
| MW50983 30 | 11/22/2020 | Injury | DEXCOM, INC. | 12/11/2020 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITORING (CGM) SYSTEM | Patient-Device Incompatibility | Local Reaction; Scarring; Burning Sensation |
| MW50981 36 | 11/23/2020 | Injury | DEXCOM, INC. | 11/27/2020 | QBJ | DEXCOM G6 SENSOR (FOR CONTINUOUS GLUCOSE MONITOR) | Difficult to Remove | Bruise/Contusion; Irritation; Swelling; Blood Loss |
| MW50977 87 | 6/1/2020 | Injury | DEXCOM, INC. | 11/10/2020 | MDS | DEXCOM G6 | Patient-Device Incompatibility | Erythema; Itching Sensation; Rash |
| MW50974 91 | 10/21/2020 | Malfunction | DEXCOM, INC. | 10/27/2020 | MDS | DEXCOM G6 SENSOR | Detachment of Device or Device Component | No Clinical Signs, Symptoms or Conditions |
| MW50974 97 | 10/25/2020 | Injury | DEXCOM, INC. | 10/27/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Adverse Event Without Identified Device or Use Problem | Skin Erosion; Skin Irritation |
| MW50974 97 | 10/25/2020 | Injury | DEXCOM, INC. | 10/27/2020 | MDS | SENSOR | Adverse Event Without Identified Device or Use Problem | Skin Erosion; Skin Irritation |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|--|--|---|
| MW50974 97 | 10/25/2020 | Injury | DEXCOM, INC. | 10/27/2020 | MDS | SENSOR | Adverse Event Without Identified Device or Use Problem | Skin Erosion; Skin Irritation |
| MW50974 69 | 10/18/2020 | Injury | DEXCOM, INC. | 10/26/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Adverse Event Without Identified Device or Use Problem | Reaction |
| MW50974 69 | 10/18/2020 | Injury | DEXCOM, INC. | 10/26/2020 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Adverse Event Without Identified Device or Use Problem | Reaction |
| MW50974 29 | | Malfunction | DEXCOM, INC. | 10/23/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITORING SYSTEM (CGM) | Product Quality Problem; Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW50973 79 | 5/1/2020 | Injury | DEXCOM, INC. | 10/21/2020 | QBJ | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Hypersensitivity/Allergic reaction; Rash; Scarring; Blister |
| MW50973 79 | 5/1/2020 | Injury | DEXCOM, INC. | 10/21/2020 | MDS | SENSORS | Adverse Event Without Identified Device or Use Problem | Hypersensitivity/Allergic reaction; Rash; Scarring; Blister |
| MW50972 43 | 10/12/2020 | Injury | DEXCOM, INC. | 10/14/2020 | MDS | DEXCOM G6 GLUCOSE SENSOR | Adverse Event Without Identified Device or Use Problem | Skin Irritation; Caustic/Chemical Burns |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|-----------------|---------------|--------------|---|---|---|
| MW50971 24 | 10/4/2020 | Injury | DEXCOM, INC. | 10/6/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Premature Discharge of Battery; Incorrect, Inadequate or Imprecise Result or Readings | Diabetic Ketoacidosis |
| MW50971 30 | 10/2/2020 | Malfunction | DEXCOM, INC. | 10/6/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Unable to Obtain Readings; Loss of Data; Inadequate User Interface | No Clinical Signs, Symptoms or Conditions |
| MW50971 30 | 10/2/2020 | Malfunction | DEXCOM, INC. | 10/6/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Unable to Obtain Readings; Loss of Data; Inadequate User Interface | No Clinical Signs, Symptoms or Conditions |
| MW50970 30 | 9/29/2020 | Injury | DEXCOM, INC. | 10/1/2020 | MDS | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Erythema; Rash |
| MW50969 25 | 9/28/2020 | Malfunction | DEXCOM, INC. | 9/28/2020 | QBJ | DEXCOM G6 CGM | Break | No Clinical Signs, Symptoms or Conditions |
| MW50969 25 | 9/28/2020 | Malfunction | DEXCOM, INC. | 9/28/2020 | --- | TRANSMITTER WITH DEXCOM CGM | Break | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|-----------------|---------------|--------------|---------------------|---|--|
| MW50968 11 | 9/16/2020 | Injury | DEXCOM, INC. | 9/22/2020 | MDS | DEXCOM CGM G6 | Retraction Problem; Activation, Positioning or Separation Problem; Material Twisted/Bent | Insufficient Information |
| MW50968 11 | 9/16/2020 | Injury | DEXCOM, INC. | 9/22/2020 | MDS | DEXCOM CGM G6 | Retraction Problem; Activation, Positioning or Separation Problem; Material Twisted/Bent | Insufficient Information |
| MW50967 39 | 8/25/2020 | Injury | DEXCOM, INC. | 9/17/2020 | QDK | DEXCOM G6 SENSOR | Output Problem; Physical Resistance/Sticki ng | Hemorrhage/Bleedi ng; Skin Discoloration |
| MW50967 39 | 8/25/2020 | Injury | DEXCOM, INC. | 9/17/2020 | QDK | DEXCOM G6 SENSOR | Output Problem; Physical Resistance/Sticki ng | Hemorrhage/Bleedi ng; Skin Discoloration |
| MW50964 99 | 9/4/2020 | Injury | DEXCOM, INC. | 9/9/2020 | MDS | DEXCOM | Product Quality Problem | Scarring; Fluid Discharge |
| MW50965 08 | 9/5/2020 | Injury | DEXCOM, INC. | 9/9/2020 | MDS | DEXCOM G6 SENSOR | Break; Product Quality Problem; Defective Device | Foreign Body In Patient |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|-----------------|---------------|--------------|------------------------------|---|---|
| MW50965 10 | | Injury | DEXCOM, INC. | 9/9/2020 | MDS | DEXCOM G6 TRANSMITTE R | Product Quality Problem | Hypersensitivity/All ergic reaction; Itching Sensation; Rash; Fluid Discharge |
| MW50965 10 | | Injury | DEXCOM, INC. | 9/9/2020 | MDS | DEXCOM G6 TRANSMITTE R | Product Quality Problem | Hypersensitivity/All ergic reaction; Itching Sensation; Rash; Fluid Discharge |
| MW50965 10 | | Injury | DEXCOM, INC. | 9/9/2020 | MDS | DEXCOM G6 TRANSMITTE R | Product Quality Problem | Hypersensitivity/All ergic reaction; Itching Sensation; Rash; Fluid Discharge |
| MW50965 10 | | Injury | DEXCOM, INC. | 9/9/2020 | MDS | DEXCOM G6 TRANSMITTE R | Product Quality Problem | Hypersensitivity/All ergic reaction; Itching Sensation; Rash; Fluid Discharge |
| MW50965 11 | 9/1/2020 | Malfunction | DEXCOM, INC. | 9/9/2020 | QBJ | DEXCOM G6 | Loss of or Failure to Bond; Device Fell | No Clinical Signs, Symptoms or Conditions |
| MW50965 11 | 9/1/2020 | Malfunction | DEXCOM, INC. | 9/9/2020 | MDS | SENSOR | Loss of or Failure to Bond; Device Fell | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|-----------------|---------------|--------------|---------------------------------------|---|--|
| MW50964 24 | 9/1/2020 | Injury | DEXCOM, INC. | 9/3/2020 | MDS | DEXCOM G6 TRANSMITTE R ADHESIVE | Patient-Device Incompatibility | Itching Sensation; Rash; Burning Sensation |
| MW50964 32 | 8/25/2020 | Malfunction | DEXCOM, INC. | 9/3/2020 | QII | DEXCOM G6 PRO | Battery Problem: Low Impedance | No Known Impact Or Consequence To Patient |
| MW50964 39 | 8/30/2020 | Injury | DEXCOM, INC. | 9/3/2020 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | Hypoglycemia |
| MW50964 40 | 8/18/2020 | Malfunction | DEXCOM, INC. | 9/3/2020 | MDS | DEXCOM G6 | Product Quality Problem | Rash; Skin Irritation |
| MW50964 40 | 8/18/2020 | Malfunction | DEXCOM, INC. | 9/3/2020 | MDS | DEXCOM G6 | Product Quality Problem | Rash; Skin Irritation |
| MW50963 72 | 8/29/2020 | Malfunction | DEXCOM, INC. | 9/1/2020 | QBJ | DEXCOM G6 | Retraction Problem | Hemorrhage/Bleedi ng |
| MW50963 16 | 1/15/2020 | Injury | DEXCOM, INC. | 8/27/2020 | MDS | DEXCOMG6 | Patient-Device Incompatibility | Unspecified Infection; Rash; Skin Irritation |
| MW50963 16 | 1/15/2020 | Injury | DEXCOM, INC. | 8/27/2020 | MDS | SENSOR | Patient-Device Incompatibility | Unspecified Infection; Rash; Skin Irritation |
| MW50962 62 | 8/22/2020 | Malfunction | DEXCOM, INC. | 8/25/2020 | QBJ | DEXCOM G6 | Communication or Transmission Problem | No Known Impact Or Consequence To Patient |
| MW50962 39 | 6/21/2020 | Injury | DEXCOM, INC. | 8/24/2020 | QBJ | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Caustic/Chemical Burns |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|-----------------|---------------|--------------|---------------------|---|---|
| MW50961 26 | 8/15/2020 | Injury | DEXCOM, INC. | 8/18/2020 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Hypersensitivity/All ergic reaction; Unspecified Infection; Itching Sensation; Skin Irritation; Fluid Discharge |
| MW50961 26 | 8/15/2020 | Injury | DEXCOM, INC. | 8/18/2020 | MDS | SENSOR | Patient-Device Incompatibility | Hypersensitivity/All ergic reaction; Unspecified Infection; Itching Sensation; Skin Irritation; Fluid Discharge |
| MW50960 94 | 8/12/2020 | Injury | DEXCOM, INC. | 8/17/2020 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Abrasion; Itching Sensation; Swelling; Discharge; Impaired Healing; Caustic/Chemical Burns |
| MW50960 27 | 8/7/2020 | Injury | DEXCOM, INC. | 8/11/2020 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Rash |
| MW50959 92 | 5/17/2020 | Injury | DEXCOM, INC. | 8/10/2020 | MDS | DEXCOM G6 CGM | Patient-Device Incompatibility | Local Reaction |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|-----------------|---------------|--------------|----------------------|--|--|
| MW50959 93 | 2/1/2020 | Injury | DEXCOM, INC. | 8/10/2020 | MDS | DEXCOM G6 SENSORS | Patient-Device Incompatibility; Detachment of Device or Device Component | Unspecified Infection; Itching Sensation; Peeling; Skin Irritation; Discharge; Discomfort |
| MW50959 68 | 7/14/2020 | Injury | DEXCOM, INC. | 8/7/2020 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Unspecified Infection; Local Reaction |
| MW50959 68 | 7/14/2020 | Injury | DEXCOM, INC. | 8/7/2020 | MDS | DEXCOM G6 | Patient-Device Incompatibility | Unspecified Infection; Local Reaction |
| MW50958 03 | 7/27/2020 | Injury | DEXCOM, INC. | 7/29/2020 | QBJ | DEXCOM SENSOR G6 | Break; Entrapment of Device | Device Embedded In Tissue or Plaque |
| MW50957 74 | 7/24/2020 | Injury | DEXCOM, INC. | 7/28/2020 | MDS | DEXCOM G6 | Patient-Device Incompatibility | Itching Sensation; Rash; Discharge; Discomfort |
| MW50957 49 | 5/27/2020 | Injury | DEXCOM | 7/27/2020 | MDS | DEXCOMG G6 SENSOR | Patient-Device Incompatibility | Rash; Scarring |
| MW50955 28 | 6/30/2020 | Injury | DEXCOM, INC. | 7/14/2020 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Purulent Discharge; Rash; Caustic/Chemical Burns |
| MW50955 28 | 6/30/2020 | Injury | DEXCOM, INC. | 7/14/2020 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Purulent Discharge; Rash; Caustic/Chemical Burns |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|--------------|---------------|--------------|-----------------------|--|---|
| MW5095528 | 6/30/2020 | Injury | DEXCOM, INC. | 7/14/2020 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Purulent Discharge; Rash; Caustic/Chemical Burns |
| MW5095513 | 6/1/2020 | Injury | DEXCOM, INC. | 7/13/2020 | QBJ | DEXCOM SENSOR | Patient-Device Incompatibility | Burn(s); Itching Sensation; Skin Irritation |
| MW5095433 | 7/4/2020 | Injury | DEXCOM, INC. | 7/8/2020 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Itching Sensation; Pain; Rash; Skin Discoloration |
| MW5095443 | 6/29/2020 | Injury | DEXCOM, INC. | 7/8/2020 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Hypersensitivity/Allergic reaction; Inflammation |
| MW5095443 | 6/29/2020 | Injury | DEXCOM, INC. | 7/8/2020 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Hypersensitivity/Allergic reaction; Inflammation |
| MW5095443 | 6/29/2020 | Injury | UNK | 7/8/2020 | FRO | DRESSING, WOUND, DRUG | Adverse Event Without Identified Device or Use Problem | Hypersensitivity/Allergic reaction; Inflammation |
| MW5095406 | 7/2/2020 | Injury | DEXCOM, INC. | 7/7/2020 | MDS | DEXCOM G6 SENSOR | Physical Resistance/Sticking | Needle Stick/Puncture |
| MW5095406 | 7/2/2020 | Injury | DEXCOM, INC. | 7/7/2020 | MDS | DEXCOM G6 SENSOR | Physical Resistance/Sticking | Needle Stick/Puncture |
| MW5095418 | 5/18/2020 | Injury | DEXCOM, INC. | 7/7/2020 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Rash; Reaction |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|--------------|---------------|--------------|---|--|--|
| MW5095418 | 5/18/2020 | Injury | DEXCOM, INC. | 7/7/2020 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Rash; Reaction |
| MW5095388 | 7/1/2020 | Injury | DEXCOM, INC. | 7/6/2020 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Rash; Blood Loss; Fluid Discharge |
| MW5095352 | 5/15/2020 | Injury | DEXCOM, INC. | 7/2/2020 | MDS | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Skin Irritation; No Code Available |
| MW5095352 | 5/15/2020 | Injury | DEXCOM, INC. | 7/2/2020 | MDS | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Skin Irritation; No Code Available |
| MW5095352 | 5/15/2020 | Injury | DEXCOM, INC. | 7/2/2020 | MDS | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Skin Irritation; No Code Available |
| MW5095352 | 5/15/2020 | Injury | DEXCOM, INC. | 7/2/2020 | MDS | DEXCOM G6 | Adverse Event Without Identified Device or Use Problem | Skin Irritation; No Code Available |
| MW5095334 | 6/25/2020 | Injury | DEXCOM, INC. | 7/1/2020 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | Patient-Device Incompatibility | Itching Sensation; Pain; Rash; Scarring; Skin Irritation |
| MW5095334 | 6/25/2020 | Injury | DEXCOM, INC. | 7/1/2020 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | Patient-Device Incompatibility | Itching Sensation; Pain; Rash; Scarring; Skin Irritation |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|--------------|---------------|--------------|--|--------------------------------|---|
| MW5095334 | 6/25/2020 | Injury | DEXCOM, INC. | 7/1/2020 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | Patient-Device Incompatibility | Itching Sensation; Pain; Rash; Scarring; Skin Irritation |
| MW5095334 | 6/25/2020 | Injury | DEXCOM, INC. | 7/1/2020 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | Patient-Device Incompatibility | Itching Sensation; Pain; Rash; Scarring; Skin Irritation |
| MW5095318 | 5/1/2020 | Injury | DEXCOM, INC. | 6/30/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITORING SYSTEM | Patient-Device Incompatibility | Reaction; Caustic/Chemical Burns |
| MW5095318 | 5/1/2020 | Injury | DEXCOM, INC. | 6/30/2020 | MDS | SENSOR | Patient-Device Incompatibility | Reaction; Caustic/Chemical Burns |
| MW5095243 | 4/20/2020 | Injury | DEXCOM, INC. | 6/25/2020 | MDS | DEXCOM G-6 | Patient-Device Incompatibility | Erythema; Itching Sensation |
| MW5095166 | 6/12/2020 | Injury | DEXCOM, INC. | 6/23/2020 | MDS | DEXCOM G6 SENSORS | Product Quality Problem | Erythema; Irritation; Itching Sensation; Swelling; Reaction |
| MW5095166 | 6/12/2020 | Injury | DEXCOM, INC. | 6/23/2020 | MDS | DEXCOM G6 SENSORS | Product Quality Problem | Erythema; Irritation; Itching Sensation; Swelling; Reaction |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|-----------------|---------------|--------------|----------------------|-----------------------------------|---|
| MW50951 66 | 6/12/2020 | Injury | DEXCOM, INC. | 6/23/2020 | MDS | DEXCOM G6 SENSORS | Product Quality Problem | Erythema; Irritation; Itching Sensation; Swelling; Reaction |
| MW50951 66 | 6/12/2020 | Injury | DEXCOM, INC. | 6/23/2020 | MDS | DEXCOM G6 SENSORS | Product Quality Problem | Erythema; Irritation; Itching Sensation; Swelling; Reaction |
| MW50951 66 | 6/12/2020 | Injury | DEXCOM, INC. | 6/23/2020 | MDS | DEXCOM G6 SENSORS | Product Quality Problem | Erythema; Irritation; Itching Sensation; Swelling; Reaction |
| MW50951 66 | 6/12/2020 | Injury | DEXCOM, INC. | 6/23/2020 | MDS | DEXCOM G6 SENSORS | Product Quality Problem | Erythema; Irritation; Itching Sensation; Swelling; Reaction |
| MW50951 70 | 5/12/2020 | Injury | DEXCOM, INC. | 6/23/2020 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Purulent Discharge; Erythema; Inflammation; Itching Sensation; Scarring; Skin Irritation; Caustic/Chemical Burns |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|--------------------------------------|--------------------------------|--|
| MW5095170 | 5/12/2020 | Injury | DEXCOM, INC. | 6/23/2020 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Purulent Discharge; Erythema; Inflammation; Itching Sensation; Scarring; Skin Irritation; Caustic/Chemical Burns |
| MW5095173 | 6/1/2020 | Injury | DEXCOM, INC. | 6/23/2020 | QBJ | DEXCOM G6 | Insufficient Information | No Known Impact Or Consequence To Patient |
| MW5095179 | 6/19/2020 | Malfunction | DEXCOM, INC. | 6/23/2020 | QBJ | DEXCOM G6 | Mechanical Problem | No Known Impact Or Consequence To Patient |
| MW5095185 | 6/16/2020 | Injury | DEXCOM, INC | 6/23/2020 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Skin Irritation |
| MW5095185 | 6/16/2020 | Injury | DEXCOM, INC | 6/23/2020 | MDS | SENSOR | Patient-Device Incompatibility | Skin Irritation |
| MW5095134 | 6/10/2020 | Injury | DEXCOM, INC. | 6/19/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Patient-Device Incompatibility | Itching Sensation; Rash; Skin Irritation |
| MW5095134 | 6/10/2020 | Injury | DEXCOM, INC. | 6/19/2020 | MDS | SENSORS | Patient-Device Incompatibility | Itching Sensation; Rash; Skin Irritation |
| MW5095134 | 6/10/2020 | Injury | DEXCOM, INC. | 6/19/2020 | MDS | SENSORS | Patient-Device Incompatibility | Itching Sensation; Rash; Skin Irritation |
| MW5095056 | 6/15/2020 | Injury | DEXCOM, INC. | 6/17/2020 | MDS | DEXCOM G6 | Patient-Device Incompatibility | Irritation; Itching Sensation; Reaction |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|--------------------------------------|--------------------------------|--|
| MW5094953 | 6/8/2020 | Malfunction | DEXCOM, INC. | 6/11/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Break | No Known Impact Or Consequence To Patient |
| MW5094953 | 6/8/2020 | Malfunction | DEXCOM, INC. | 6/11/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Break | No Known Impact Or Consequence To Patient |
| MW5094889 | 6/4/2020 | Injury | DEXCOM INC. | 6/8/2020 | QBJ | DEXCOM G6 CGM | Patient-Device Incompatibility | Pain; Skin Irritation; No Code Available |
| MW5094824 | 6/1/2020 | Injury | DEXCOM, INC. | 6/3/2020 | QBJ | DEXCOM | Patient-Device Incompatibility | Rash |
| MW5094804 | 5/29/2020 | Injury | DEXCOM, INC. | 6/2/2020 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Rash; Skin Irritation |
| MW5094804 | 5/29/2020 | Injury | DEXCOM, INC. | 6/2/2020 | MDS | DEXCOM SENSOR | Patient-Device Incompatibility | Rash; Skin Irritation |
| MW5094804 | 5/29/2020 | Injury | DEXCOM, INC. | 6/2/2020 | MDS | DEXCOM SENSOR | Patient-Device Incompatibility | Rash; Skin Irritation |
| MW5094804 | 5/29/2020 | Injury | DEXCOM, INC. | 6/2/2020 | MDS | DEXCOM SENSOR | Patient-Device Incompatibility | Rash; Skin Irritation |
| MW5094808 | 5/29/2020 | Injury | DEXCOM, INC. | 6/2/2020 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Itching Sensation; Rash; Local Reaction; Skin Irritation; Blood Loss |
| MW5094808 | 5/29/2020 | Injury | DEXCOM, INC. | 6/2/2020 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Itching Sensation; Rash; Local Reaction; Skin Irritation; Blood Loss |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|---|--------------------------------|---|
| MW5094718 | 5/22/2020 | Injury | DEXCOM, INC. | 5/27/2020 | MDS | DEXCOM CGM | Product Quality Problem | Erythema; Hemorrhage/Bleeding; Pain; Swelling; Complaint, Ill-Defined |
| MW5094718 | 5/22/2020 | Injury | DEXCOM, INC. | 5/27/2020 | MDS | DEXCOM CGM | Product Quality Problem | Erythema; Hemorrhage/Bleeding; Pain; Swelling; Complaint, Ill-Defined |
| MW5094718 | 5/22/2020 | Injury | DEXCOM, INC. | 5/27/2020 | MDS | DEXCOM CGM | Product Quality Problem | Erythema; Hemorrhage/Bleeding; Pain; Swelling; Complaint, Ill-Defined |
| MW5094720 | 5/25/2020 | Malfunction | DEXCOM, INC. | 5/27/2020 | QBJ | DEXCOM G6 SENSOR | Separation Failure | No Known Impact Or Consequence To Patient |
| MW5094721 | 1/21/2020 | Injury | DEXCOM, INC. | 5/27/2020 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | Patient-Device Incompatibility | Itching Sensation; Rash |
| MW5094728 | 5/19/2020 | Injury | DEXCOM, INC. | 5/27/2020 | MDS | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | Patient-Device Incompatibility | Irritation; Itching Sensation; Burning Sensation |
| MW5094578 | 3/1/2020 | Injury | DEXCOM, INC. | 5/19/2020 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Reaction |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|--------------|---------------|--------------|------------------|--|---|
| MW5094539 | 5/1/2020 | Injury | DEXCOM, INC. | 5/18/2020 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Impaired Healing; Caustic/Chemical Burns |
| MW5094500 | 5/2/2020 | Injury | DEXCOM, INC. | 5/14/2020 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Hypersensitivity/Allergic reaction; Itching Sensation; Rash |
| MW5094500 | 5/2/2020 | Injury | DEXCOM, INC. | 5/14/2020 | MDS | DEXCOM G6 SENSOR | Adverse Event Without Identified Device or Use Problem | Hypersensitivity/Allergic reaction; Itching Sensation; Rash |
| MW5094502 | 2/20/2020 | Injury | DEXCOM, INC. | 5/14/2020 | MDS | DEXCOM G6 SENSOR | Patient-Device Incompatibility | Itching Sensation; Pain; Rash; Scar Tissue; Discharge; Caustic/Chemical Burns |
| MW5094506 | 4/1/2020 | Injury | DEXCOM, INC. | 5/14/2020 | QBJ | DEXCOM G6 | Patient-Device Incompatibility | Reaction |
| MW5094464 | 5/8/2020 | Injury | DEXCOM, INC. | 5/12/2020 | QDK | DEXCOM G6 | Patient-Device Incompatibility | Rash; Blood Loss |
| MW5094465 | 4/6/2020 | Injury | DEXCOM, INC. | 5/12/2020 | QBJ | DEXCOM G6 CGM | Patient-Device Incompatibility | Skin Irritation; Blood Loss |
| MW5094390 | 5/1/2020 | Injury | DEXCOM, INC. | 5/5/2020 | QBJ | DEXCOM G6 | Product Quality Problem | Erythema; Itching Sensation; Rash; Reaction |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|---|---|---|
| MW50943 90 | 5/1/2020 | Injury | DEXCOM, INC. | 5/5/2020 | QBJ | DEXCOM G6 | Product Quality Problem | Erythema; Itching Sensation; Rash; Reaction |
| MW50943 90 | 5/1/2020 | Injury | DEXCOM, INC. | 5/5/2020 | QBJ | DEXCOM G6 | Product Quality Problem | Erythema; Itching Sensation; Rash; Reaction |
| MW50943 91 | 5/1/2020 | Injury | DEXCOM, INC. | 5/5/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | Product Quality Problem | Rash; Swelling; Reaction |
| MW50942 35 | 4/19/2020 | Injury | DEXCOM, INC. | 4/22/2020 | QBJ | DEXCOM G6 CONSTANT GLUCOSE MONITOR | Difficult to Insert; Device Dislodged or Dislocated; Output Problem | Pain |
| MW50939 25 | 3/24/2020 | Malfunction | DEXCOM, INC. | 3/25/2020 | QBJ | DEXCOM G6 CGM SENSOR | Device Alarm System; Battery Problem | Hyperglycemia |
| MW50939 25 | 3/24/2020 | Malfunction | DEXCOM, INC. | 3/25/2020 | QBJ | DEXCOM G6 CGM SENSOR | Device Alarm System; Battery Problem | Hyperglycemia |
| MW50939 25 | 3/24/2020 | Malfunction | DEXCOM, INC. | 3/25/2020 | QBJ | DEXCOM G6 CGM SENSOR | Device Alarm System; Battery Problem | Hyperglycemia |
| MW50934 87 | 12/28/2019 | Injury | DEXCOM, INC. | 3/2/2020 | QBJ | DEXCOM G6 CGM | Retraction Problem; Manufacturing, Packaging or Shipping Problem | Hemorrhage/Bleeding; Pain |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|--|--|---|
| MW5093487 | 12/28/2019 | Injury | DEXCOM, INC. | 3/2/2020 | QBJ | DEXCOM G6 CGM | Retraction Problem; Manufacturing, Packaging or Shipping Problem | Hemorrhage/Bleeding; Pain |
| MW5093247 | 2/12/2020 | Injury | DEXCOM, INC. | 2/24/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Device Fell | No Known Impact Or Consequence To Patient |
| MW5093055 | 2/13/2020 | Injury | DEXCOM, INC. | 2/18/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITORING DEVICE | Defective Device | No Known Impact Or Consequence To Patient |
| MW5092936 | 2/7/2020 | Injury | DEXCOM, INC. | 2/11/2020 | QBJ | DEXCOM G6 | Difficult to Remove | No Known Impact Or Consequence To Patient |
| MW5092767 | 1/11/2020 | Malfunction | DEXCOM, INC. | 2/4/2020 | QBJ | DEXCOM G6 | Connection Problem | No Known Impact Or Consequence To Patient |
| MW5092601 | 1/24/2020 | Malfunction | DEXCOM, INC. | 1/28/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Electrical /Electronic Property Problem; Communication or Transmission Problem; Device Sensing Problem | No Known Impact Or Consequence To Patient |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|---|--|---|
| MW5092601 | 1/24/2020 | Malfunction | DEXCOM, INC. | 1/28/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Electrical /Electronic Property Problem; Communication or Transmission Problem; Device Sensing Problem | No Known Impact Or Consequence To Patient |
| MW5092601 | 1/24/2020 | Malfunction | DEXCOM, INC. | 1/28/2020 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Electrical /Electronic Property Problem; Communication or Transmission Problem; Device Sensing Problem | No Known Impact Or Consequence To Patient |
| MW5092555 | 1/24/2020 | Injury | DEXCOM, INC. | 1/27/2020 | QBJ | DEXCOM G6 | Mechanical Problem | No Known Impact Or Consequence To Patient |
| MW5092455 | 1/17/2020 | Injury | DEXCOM, INC. | 1/22/2020 | QDK | DEXCOM G6CONTINUOUS GLUCOSE MONITORING DEVICE | Computer Software Problem; Loss of Data | Hypoglycemia |
| MW5092457 | 1/11/2020 | Injury | DEXCOM, INC. | 1/22/2020 | QDK | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Retraction Problem; Physical Resistance/Sticking | Pain |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|--------------------------------------|---|---|
| MW5092457 | 1/11/2020 | Injury | DEXCOM, INC. | 1/22/2020 | QDK | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Retraction Problem; Physical Resistance/Sticking | Pain |
| MW5092457 | 1/11/2020 | Injury | DEXCOM, INC. | 1/22/2020 | QDK | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Retraction Problem; Physical Resistance/Sticking | Pain |
| MW5091938 | 12/26/2019 | Malfunction | DEXCOM, INC. | 12/30/2019 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Unable to Obtain Readings | No Known Impact Or Consequence To Patient |
| MW5091396 | 11/30/2019 | Injury | DEXCOM, INC. | 12/3/2019 | QBJ | DEXCOM G6 | Disconnection; Incorrect, Inadequate or Imprecise Result or Readings; Low Test Results; Device Displays Incorrect Message | Hyperglycemia |
| MW5091397 | 12/1/2019 | Injury | DEXCOM, INC. | 12/3/2019 | QBJ | DEXCOM G6 | Wireless Communication Problem | No Known Impact Or Consequence To Patient |
| MW5091440 | 11/30/2019 | Malfunction | DEXCOM, INC. | 12/3/2019 | QBJ | DEXCOM G6 | Application Program Problem | Hypoglycemia; Cognitive Changes |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|------------|-----------------|---------------|--------------|------------|---|-----------------|
| MW50905 42 | 3/10/2011 | Injury | DEXCOM, INC. | 10/21/2019 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Defective Component; High Readings | Hypoglycemia |
| MW50905 42 | 3/10/2011 | Injury | DEXCOM, INC. | 10/21/2019 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Defective Component; High Readings | Hypoglycemia |
| MW50905 42 | 3/10/2011 | Injury | DEXCOM, INC. | 10/21/2019 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Defective Component; High Readings | Hypoglycemia |
| MW50905 42 | 3/10/2011 | Injury | DEXCOM, INC. | 10/21/2019 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Defective Component; High Readings | Hypoglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|--------------------|---|---|
| MW5090542 | 3/10/2011 | Injury | DEXCOM, INC. | 10/21/2019 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Defective Component; High Readings | Hypoglycemia |
| MW5090423 | 10/13/2019 | Malfunction | DEXCOM, INC. | 10/15/2019 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | No Known Impact Or Consequence To Patient |
| MW5089729 | 9/9/2019 | Malfunction | DEXCOM, INC. | 9/11/2019 | QBJ | DEXCOM TRANSMITTER | Electromagnetic Interference | No Known Impact Or Consequence To Patient |
| MW5089688 | 9/8/2019 | Injury | DEXCOM, INC. | 9/10/2019 | QBJ | DEXCOM G6 SENSOR | Defective Component | No Known Impact Or Consequence To Patient |
| MW5089421 | 8/26/2019 | Injury | DEXCOM, INC. | 8/28/2019 | QBJ | DEXCOM G6 | High Readings | No Known Impact Or Consequence To Patient |
| MW5089191 | 8/15/2019 | Malfunction | DEXCOM, INC. | 8/19/2019 | QBJ | DEXCOM G6 CGM | Incorrect, Inadequate or Imprecise Result or Readings | No Known Impact Or Consequence To Patient |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|----------------------------------|---------------|--------------|-------------------------|---|---|
| MW50887 68 | 10/29/2019 | Malfunction | DEXCOM, INC. | 8/5/2019 | QBJ | DEXCOM G6 CGM SENSOR | Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings | No Known Impact Or Consequence To Patient |
| MW50887 68 | 10/29/2019 | Malfunction | TANDEM DIABETES CARE, INC. | 8/5/2019 | PQF | T:SLIMX2 | Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings | No Known Impact Or Consequence To Patient |
| MW50887 68 | 10/29/2019 | Malfunction | DEXCOM INC. | 8/5/2019 | --- | DEXCOM G6 CGM SENSOR | Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings | No Known Impact Or Consequence To Patient |
| MW50887 68 | 10/29/2019 | Malfunction | DEXCOM, INC. | 8/5/2019 | --- | DEXCOM G6 CGM SENSOR | Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings | No Known Impact Or Consequence To Patient |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|-----------------|---------------|--------------|---|---|---|
| MW50887 68 | 10/29/2019 | Malfunction | DEXCOM, INC. | 8/5/2019 | --- | DEXCOM G6 CGM SENSOR | Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings | No Known Impact Or Consequence To Patient |
| MW50887 68 | 10/29/2019 | Malfunction | DEXCOM, INC. | 8/5/2019 | --- | DEXCOM G6 CGM SENSOR | Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings | No Known Impact Or Consequence To Patient |
| MW50887 68 | 10/29/2019 | Malfunction | DEXCOM, INC. | 8/5/2019 | --- | DEXCOM G6 CGM SENSOR | Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings | No Known Impact Or Consequence To Patient |
| MW50885 41 | 7/25/2019 | Injury | DEXCOM, INC. | 7/29/2019 | QBJ | DEXCOM G6 SENSOR | Break; Sensing Intermittently | No Known Impact Or Consequence To Patient |
| MW50855 74 | 3/31/2019 | Injury | DEXCOM, INC | 4/3/2019 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Defective Device | Pain; Blood Loss |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------|-------------|--------------|---------------|--------------|--------------------------------------|---|---|
| MW5085574 | 3/31/2019 | Injury | DEXCOM, INC | 4/3/2019 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Defective Device | Pain; Blood Loss |
| MW5085574 | 3/31/2019 | Injury | DEXCOM, INC | 4/3/2019 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Defective Device | Pain; Blood Loss |
| MW5084988 | 3/14/2019 | Malfunction | DEXCOM INC. | 3/18/2019 | QDK | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Detachment of Device or Device Component; Material Deformation | No Known Impact Or Consequence To Patient |
| MW5084487 | 2/18/2019 | NA | DEXCOM INC. | 2/27/2019 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Display or Visual Feedback Problem; Therapeutic or Diagnostic Output Failure | Hypoglycemia |
| MW5083130 | 1/10/2019 | Injury | DEXCOM, INC. | 1/15/2019 | QBJ | DEXCOM G6 | False Alarm; Incorrect, Inadequate or Imprecise Result or Readings; Protective Measures Problem | Hypoglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-----------------------|-------------|-----------------|-----------------------|--------------|---|--|---|
| MW50831 30 | 1/10/2019 | Injury | DEXCOM, INC. | 1/15/2019 | QBJ | DEXCOM G6 | False Alarm; Incorrect, Inadequate or Imprecise Result or Readings; Protective Measures Problem | Hypoglycemia |
| MW50831 30 | 1/10/2019 | Injury | DEXCOM, INC. | 1/15/2019 | QBJ | DEXCOM G6 | False Alarm; Incorrect, Inadequate or Imprecise Result or Readings; Protective Measures Problem | Hypoglycemia |
| MW50831 30 | 1/10/2019 | Injury | DEXCOM, INC. | 1/15/2019 | QBJ | DEXCOM G6 | False Alarm; Incorrect, Inadequate or Imprecise Result or Readings; Protective Measures Problem | Hypoglycemia |
| MW50829 36 | 1/4/2019 | Malfunction | DEXCOM INC. | 1/7/2019 | QDK | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR | Material Integrity Problem | No Known Impact Or Consequence To Patient |
| MW51617 50 | 2024/10/19 4:00:00 | Malfunction | DEXCOM, INC. | 2024/10/28 4:00:00 | QBJ | DEXCOM G7 | | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-----------------------|-------------|--------------|-----------------------|--------------|--------------------------------------|----------------|---|
| MW5160849 | 2024/06/20 4:00:00 | Malfunction | DEXCOM, INC | 2024/10/08 4:00:00 | QBJ | DEXCOM G7 | | Device Embedded In Tissue or Plaque |
| MW5160431 | 2024/07/26 4:00:00 | Malfunction | DEXCOM, INC. | 2024/09/30 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE MONITOR | | No Clinical Signs, Symptoms or Conditions |
| MW5160432 | 2024/07/31 4:00:00 | Malfunction | DEXCOM, INC. | 2024/09/30 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE MONITOR | | No Clinical Signs, Symptoms or Conditions |
| MW5160433 | 2024/08/10 4:00:00 | Malfunction | DEXCOM, INC. | 2024/09/30 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE MONITOR | | No Clinical Signs, Symptoms or Conditions |
| MW5160434 | 2024/09/22 4:00:00 | Malfunction | DEXCOM, INC. | 2024/09/30 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE MONITOR | | No Clinical Signs, Symptoms or Conditions |
| MW5159660 | 2024/09/14 4:00:00 | Malfunction | DEXCOM, INC. | 2024/09/17 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE SENSOR | | Insufficient Information |
| MW5159661 | 2024/09/14 4:00:00 | Malfunction | DEXCOM, INC. | 2024/09/17 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5159662 | 2024/09/14 4:00:00 | Malfunction | DEXCOM, INC. | 2024/09/17 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5159553 | 2024/09/06 4:00:00 | Injury | DEXCOM, INC. | 2024/09/11 4:00:00 | QBJ | DEXCOM G7 | | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-----------------------|-------------|--------------|-----------------------|--------------|--------------------------------------|----------------|---|
| MW5159459 | 2024/08/23 4:00:00 | Injury | DEXCOM, INC. | 2024/09/09 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE MONITOR | | Insufficient Information |
| MW5159232 | 2024/08/13 4:00:00 | Malfunction | DEXCOM, INC. | 2024/09/05 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE MONITOR | | No Clinical Signs, Symptoms or Conditions |
| MW5159233 | 2024/08/13 4:00:00 | Malfunction | DEXCOM, INC. | 2024/09/05 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE MONITOR | | No Clinical Signs, Symptoms or Conditions |
| MW5159234 | 2024/08/13 4:00:00 | Malfunction | DEXCOM, INC. | 2024/09/05 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE MONITOR | | No Clinical Signs, Symptoms or Conditions |
| MW5159270 | 2024/09/03 4:00:00 | Malfunction | DEXCOM, INC. | 2024/09/05 4:00:00 | QBJ | DEXCOM G7 | | Insufficient Information |
| MW5158720 | 2024/08/16 4:00:00 | Injury | DEXCOM, INC. | 2024/08/21 4:00:00 | QBJ | DEXCOM G7 | | Insufficient Information |
| MW5158742 | 2024/08/09 4:00:00 | Malfunction | DEXCOM, INC. | 2024/08/21 4:00:00 | QBJ | DEXCOM G7 | | Hypoglycemia |
| MW5158743 | 2024/08/09 4:00:00 | Malfunction | DEXCOM, INC. | 2024/08/21 4:00:00 | QBJ | DEXCOM G7 | | Hypoglycemia |
| MW5158744 | 2024/08/09 4:00:00 | Malfunction | DEXCOM, INC. | 2024/08/21 4:00:00 | QBJ | DEXCOM G7 | | Hypoglycemia |
| MW5157923 | 2024/07/27 4:00:00 | Injury | DEXCOM, INC. | 2024/08/01 4:00:00 | QBJ | DEXCOM G7 | | No Clinical Signs, Symptoms or Conditions |
| MW5157752 | 2024/04/17 4:00:00 | Injury | DEXCOM, INC. | 2024/07/26 4:00:00 | QBJ | DEXCOM G7 | | No Clinical Signs, Symptoms or Conditions |
| MW5157005 | 2024/06/30 4:00:00 | Malfunction | DEXCOM INC | 2024/07/03 4:00:00 | QBJ | DEXCOM G7 SENSOR | | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-----------------------|-------------|--------------|-----------------------|--------------|-------------------|----------------|---|
| MW5157006 | 2024/06/30 4:00:00 | Malfunction | DEXCOM INC. | 2024/07/03 4:00:00 | QBJ | DEXCOM G7 SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5157007 | 2024/06/30 4:00:00 | Malfunction | DEXCOM INC. | 2024/07/03 4:00:00 | QBJ | DEXCOM G7 SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5157008 | 2024/06/30 4:00:00 | Malfunction | DEXCOM INC. | 2024/07/03 4:00:00 | QBJ | DEXCOM G7 SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5157039 | 2024/06/29 4:00:00 | Injury | DEXCOM, INC. | 2024/07/03 4:00:00 | QBJ | DEXCOM G7 | | Hypoglycemia |
| MW5156887 | 2024/06/27 4:00:00 | Malfunction | DEXCOM, INC. | 2024/07/01 4:00:00 | QBJ | DEXCOM G7 SENSORS | | Insufficient Information |
| MW5156888 | 2024/06/27 4:00:00 | Malfunction | DEXCOM, INC. | 2024/07/01 4:00:00 | QBJ | DEXCOM G7 SENSORS | | Insufficient Information |
| MW5156889 | 2024/06/27 4:00:00 | Malfunction | DEXCOM, INC. | 2024/07/01 4:00:00 | QBJ | DEXCOM G7 SENSORS | | Insufficient Information |
| MW5156912 | 2024/06/26 4:00:00 | Malfunction | DEXCOM, INC. | 2024/07/01 4:00:00 | QBJ | DEXCOM G7 | | No Clinical Signs, Symptoms or Conditions |
| MW5156744 | 2024/06/18 4:00:00 | Injury | DEXCOM, INC. | 2024/06/27 4:00:00 | QBJ | DEXCOM G7 | | Hemorrhage/Bleeding |
| MW5156745 | 2024/06/18 4:00:00 | Injury | DEXCOM, INC. | 2024/06/27 4:00:00 | QBJ | DEXCOM G7 | | Hemorrhage/Bleeding |
| MW5156699 | 2024/06/03 4:00:00 | Injury | DEXCOM, INC. | 2024/06/26 4:00:00 | QBJ | DEXCOM G7 | | Hypoglycemia |
| MW5156572 | 2024/06/21 4:00:00 | Injury | DEXCOM, INC. | 2024/06/25 4:00:00 | QBJ | DEXCOM G7 CGM | | Hyperglycemia |
| MW5156573 | 2024/06/21 4:00:00 | Injury | DEXCOM, INC. | 2024/06/25 4:00:00 | QBJ | DEXCOM G7 CGM | | Hyperglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|--------------------|-------------|--------------|--------------------|--------------|--|----------------|---|
| MW5156574 | 2024/06/21 4:00:00 | Injury | DEXCOM, INC. | 2024/06/25 4:00:00 | QBJ | DEXCOM G7 CGM | | Hyperglycemia |
| MW5156575 | 2024/06/21 4:00:00 | Injury | DEXCOM, INC. | 2024/06/25 4:00:00 | QBJ | DEXCOM G7 CGM | | Hyperglycemia |
| MW5156576 | 2024/06/21 4:00:00 | Injury | DEXCOM, INC. | 2024/06/25 4:00:00 | QBJ | DEXCOM G7 CGM | | Hyperglycemia |
| MW5153370 | 2024/03/13 4:00:00 | Injury | DEXCOM, INC. | 2024/03/28 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE MONITORING SYSTEM | | No Clinical Signs, Symptoms or Conditions |
| MW5152822 | 2024/03/11 4:00:00 | Malfunction | DEXCOM, INC. | 2024/03/13 4:00:00 | QBJ | DEXCOM G7 | | No Clinical Signs, Symptoms or Conditions |
| MW5152286 | 2024/02/23 5:00:00 | Injury | DEXCOM, INC. | 2024/03/04 5:00:00 | QBJ | DEXCOM G7 SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5152287 | 2024/02/23 5:00:00 | Malfunction | DEXCOM, INC. | 2024/03/04 5:00:00 | QBJ | DEXCOM G7 SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5152288 | 2024/02/23 5:00:00 | Malfunction | DEXCOM, INC. | 2024/03/04 5:00:00 | QBJ | DEXCOM G7 SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5152289 | 2024/02/23 5:00:00 | Malfunction | DEXCOM, INC. | 2024/03/04 5:00:00 | QBJ | DEXCOM G7 SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5152290 | 2024/02/23 5:00:00 | Malfunction | DEXCOM, INC. | 2024/03/04 5:00:00 | QBJ | DEXCOM G7 SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5152291 | 2024/02/23 5:00:00 | Malfunction | DEXCOM, INC. | 2024/03/04 5:00:00 | QBJ | DEXCOM G7 SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5152177 | 2024/02/01 5:00:00 | Injury | DEXCOM, INC. | 2024/02/27 5:00:00 | QBJ | DEXCOM G7 SENSOR | | Rash; Blister |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|--------------------|-------------|--------------|--------------------|--------------|--------------------------------------|----------------|---|
| MW5151824 | 2024/02/15 5:00:00 | Injury | DEXCOM, INC. | 2024/02/21 5:00:00 | QBJ | DEXCOM G7 | | Hypoglycemia |
| MW5151008 | 2024/01/30 5:00:00 | Malfunction | DEXCOM, INC. | 2024/02/02 5:00:00 | QBJ | DEXCOM G7 SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5151009 | 2024/01/30 5:00:00 | Malfunction | DEXCOM, INC. | 2024/02/02 5:00:00 | QBJ | DEXCOM G7 SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5150844 | 2024/01/13 5:00:00 | Injury | DEXCOM, INC. | 2024/01/26 5:00:00 | QBJ | DEXCOM G7 | | Insufficient Information |
| MW5150787 | 2024/01/17 5:00:00 | Injury | DEXCOM, INC. | 2024/01/25 5:00:00 | QBJ | DEXCOM G7 | | Itching Sensation; Skin Inflammation/Irritation |
| MW5150679 | 2024/01/11 5:00:00 | Injury | DEXCOM, INC. | 2024/01/23 5:00:00 | QBJ | DEXCOM | | Hyperglycemia; Diabetic Ketoacidosis |
| MW5150080 | 2023/12/30 5:00:00 | Malfunction | DEXCOM, INC. | 2024/01/09 5:00:00 | QBJ | DEXCOM G7 | | Sleep Dysfunction |
| MW5149773 | 2023/12/27 5:00:00 | Injury | DEXCOM INC. | 2024/01/03 5:00:00 | QBJ | DEXCOM G7 | | No Clinical Signs, Symptoms or Conditions |
| MW5149365 | 2023/12/08 5:00:00 | Injury | DEXCOM, INC. | 2023/12/18 5:00:00 | QBJ | DEXCOM G7 | | Bacterial Infection; Septic Shock |
| MW5149167 | 2023/12/07 5:00:00 | Injury | DEXCOM, INC. | 2023/12/12 5:00:00 | QBJ | DEXCOM G7 | | Needle Stick/Puncture |
| MW5149168 | 2023/12/07 5:00:00 | Injury | DEXCOM, INC. | 2023/12/12 5:00:00 | QBJ | DEXCOM G7 | | Needle Stick/Puncture |
| MW5147980 | 2023/10/18 4:00:00 | Injury | DEXCOM, INC. | 2023/11/08 5:00:00 | QBJ | DEXCOM G7 CGM SYSTEM | | Anxiety; Sleep Dysfunction |
| MW5146655 | 2023/09/29 4:00:00 | Injury | DEXCOM, INC. | 2023/10/05 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE MONITOR | | Foreign Body In Patient |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|--------------------|-------------|--------------|--------------------|--------------|---|--|---|
| MW5146255 | 2023/09/23 4:00:00 | Malfunction | DEXCOM, INC. | 2023/09/27 4:00:00 | QBJ | DEXCOM G7 SENSOR | | Insufficient Information |
| MW5145972 | 2023/09/17 4:00:00 | Malfunction | DEXCOM, INC. | 2023/09/20 4:00:00 | QBJ | DEXCOM G7 CONTINUOUS GLUCOSE MONITORING | | No Clinical Signs, Symptoms or Conditions |
| MW5145611 | 2023/08/04 4:00:00 | Malfunction | DEXCOM INC. | 2023/09/13 4:00:00 | QBJ | DEXCOM G7 CGM | | No Clinical Signs, Symptoms or Conditions |
| MW5145617 | 2023/08/22 4:00:00 | Malfunction | DEXCOM, INC. | 2023/09/13 4:00:00 | QBJ | DEXCOM G7 CGM | Incorrect, Inadequate or Imprecise Result or Readings; Defective Device | No Clinical Signs, Symptoms or Conditions |
| MW5145618 | 2023/08/30 4:00:00 | Malfunction | DEXCOM, INC. | 2023/09/13 4:00:00 | QBJ | DEXCOM G7 CGM | Incorrect, Inadequate or Imprecise Result or Readings; Calibration Problem; Device Sensing Problem; Data Problem | No Clinical Signs, Symptoms or Conditions |
| MW5145620 | 2023/08/15 4:00:00 | Malfunction | DEXCOM, INC. | 2023/09/13 4:00:00 | QBJ | DEXCOM G7 CGM | Low Readings; Device Sensing Problem; Data Problem | No Clinical Signs, Symptoms or Conditions |
| MW5145621 | 2023/08/30 4:00:00 | Malfunction | DEXCOM INC. | 2023/09/13 4:00:00 | QBJ | DEXCOM G7 CGM | | No Clinical Signs, Symptoms or Conditions |
| MW5145622 | 2023/09/07 4:00:00 | Malfunction | DEXCOM INC. | 2023/09/13 4:00:00 | QBJ | DEXCOM G7 CGM | | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-----------------------|-------------|--------------|-----------------------|--------------|-------------------|--|---|
| MW5145623 | 2023/09/07 4:00:00 | Malfunction | DEXCOM INC. | 2023/09/13 4:00:00 | QBJ | DEXCOM G7 CGM | | No Clinical Signs, Symptoms or Conditions |
| MW5144829 | 2023/08/17 4:00:00 | Injury | DEXCOM, INC. | 2023/08/21 4:00:00 | QBJ | DEXCOM G7 | | No Clinical Signs, Symptoms or Conditions |
| MW5118742 | 2023/06/20 4:00:00 | Injury | DEXCOM, INC. | 2023/06/23 4:00:00 | QBJ | DEXCOM G7 READER | Defective Alarm; Product Quality Problem; Device Sensing Problem; Wireless Communication Problem | Hypoglycemia |
| MW5118743 | 2023/06/20 4:00:00 | Injury | DEXCOM, INC. | 2023/06/23 4:00:00 | MDS | DEXCOM G7 SENSORS | Defective Alarm; Product Quality Problem; Device Sensing Problem; Wireless Communication Problem | Hypoglycemia |
| MW5118744 | 2023/06/20 4:00:00 | Injury | DEXCOM, INC. | 2023/06/23 4:00:00 | MDS | DEXCOM G7 SENSORS | Defective Alarm; Product Quality Problem; Device Sensing Problem; Wireless Communication Problem | Hypoglycemia |
| MW5118745 | 2023/06/20 4:00:00 | Injury | DEXCOM, INC. | 2023/06/23 4:00:00 | MDS | DEXCOM G7 SENSORS | No Audible Alarm; Product Quality Problem; Activation Failure; Wireless Communication Problem | Hypoglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-----------------------|-------------|-----------------|-----------------------|--------------|----------------------|--|---|
| MW51187 46 | 2023/06/20 4:00:00 | Injury | DEXCOM, INC. | 2023/06/23 4:00:00 | MDS | DEXCOM G7 SENSORS | Defective Alarm; Product Quality Problem; Device Sensing Problem; Wireless Communication Problem | Hypoglycemia |
| MW51187 47 | 2023/06/20 4:00:00 | Injury | DEXCOM, INC. | 2023/06/23 4:00:00 | MDS | DEXCOM G7 SENSORS | No Audible Alarm; Product Quality Problem; Incorrect, Inadequate or Imprecise Result or Readings; Activation Failure; Wireless Communication Problem | Hypoglycemia |
| MW51187 48 | 2023/06/20 4:00:00 | Injury | DEXCOM, INC. | 2023/06/23 4:00:00 | MDS | DEXCOM G7 SENSORS | Defective Alarm; Product Quality Problem; Device Sensing Problem; Wireless Communication Problem | Hypoglycemia |
| MW51176 16 | 2023/05/01 4:00:00 | Malfunction | DEXCOM, INC. | 2023/05/17 4:00:00 | QBJ | DEXCOMG7 | Use of Device Problem; Difficult to Open or Remove Packaging Material | No Clinical Signs, Symptoms or Conditions |
| MW51176 17 | 2023/05/01 4:00:00 | Malfunction | DEXCOM, INC. | 2023/05/17 4:00:00 | QBJ | DEXCOMG7 | Use of Device Problem | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|----------------------------|-----------------------|-------------|--------------------------------|-----------------------|--------------|--|---|--|
| MW51176 18 | 2023/05/01 4:00:00 | Malfunction | DEXCOM, INC. | 2023/05/17 4:00:00 | QBJ | DEXCOMG7 | Product Quality Problem; Defective Component | No Clinical Signs, Symptoms or Conditions |
| MW51171 42 | 2023/04/26 4:00:00 | Injury | DEXCOM, INC. | 2023/04/28 4:00:00 | QBJ | DEXCOM G7 | Use of Device Problem; Communication or Transmission Problem | Hypoglycemia |
| MW51171 43 | 2023/04/26 4:00:00 | Injury | DEXCOM, INC. | 2023/04/28 4:00:00 | MDS | DEXCOM G7 SENSOR | Use of Device Problem; Communication or Transmission Problem | Hypoglycemia |
| MW51594 27 | 8/30/2024 4:00 | Malfunction | DEXCOM, INC. | 9/9/2024 4:00 | SAF | DEXCOM STELO CGM | | No Clinical Signs, Symptoms or Conditions |
| 2954323- 2024- 43124 | 2024/10/30 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/31 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Low Readings | Loss of consciousness; Confusion/ Disorientation |
| 2954323- 2024- 43147 | 2024/10/23 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/31 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia; Loss of consciousness |
| 2954323- 2024- 42804 | 2024/10/15 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/30 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323- 2024- 42665 | 2024/10/16 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/29 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|--------------------|--------------------|-------------|--------------------------|--------------------|--------------|-------------------------------|--|---|
| 2954323-2024-42685 | 2024/10/21 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/29 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-42774 | 2024/10/22 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/29 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia; Loss of consciousness |
| 2954323-2024-42470 | 2024/10/21 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/28 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-42410 | 2024/09/28 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/26 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Device Displays Incorrect Message | Hypoglycemia |
| 2954323-2024-42101 | 2024/10/15 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/24 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-41953 | 2024/10/09 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/23 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Diabetic Ketoacidosis |
| 2954323-2024-41652 | 2024/09/23 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/22 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | High Readings | Hypoglycemia; Loss of consciousness |
| 2954323-2024-41718 | 2024/10/18 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/22 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|--------------------|--------------------|-------------|--------------------------|--------------------|--------------|-------------------------------|--|---|
| 2954323-2024-41720 | 2024/10/02 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/22 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-41726 | 2024/10/10 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/22 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-41738 | 2024/10/09 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/22 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-41397 | 2024/10/08 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/21 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-41590 | 2024/10/18 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/21 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Loss of consciousness |
| 2954323-2024-41326 | 2024/10/17 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/18 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings | Loss of consciousness |
| 2954323-2024-40975 | 2024/10/11 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/17 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-40986 | 2024/10/03 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/17 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|--------------------|--------------------|-------------|--------------------------|--------------------|--------------|-------------------------------|--|---|
| 2954323-2024-41022 | 2024/10/11 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/17 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-41105 | 2024/10/14 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/17 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia |
| 2954323-2024-41128 | 2024/10/08 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/17 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Low Readings | Loss of consciousness |
| 2954323-2024-41131 | 2024/09/27 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/17 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Headache |
| 2954323-2024-40443 | 2024/09/30 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/14 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-40379 | 2024/09/30 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/13 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Low Readings | Hyperglycemia |
| 2954323-2024-40091 | 2024/10/06 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/09 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hyperglycemia |
| 2954323-2024-39716 | 2024/09/13 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/08 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|--------------------|--------------------|-------------|--------------------------|--------------------|--------------|-------------------------------|--|---|
| 2954323-2024-39781 | 2024/09/11 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/08 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia |
| 2954323-2024-39850 | 2024/09/12 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/08 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Detachment of Device or Device Component | Hyperglycemia |
| 2954323-2024-39864 | 2024/09/11 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/08 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Device Displays Incorrect Message | Hypoglycemia |
| 2954323-2024-39865 | 2024/09/10 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/08 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Device Displays Incorrect Message | Hypoglycemia |
| 2954323-2024-39621 | 2024/09/09 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/07 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Low Readings | Hyperglycemia; Loss of consciousness; Diaphoresis |
| 2954323-2024-39385 | 2024/09/11 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/04 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-38962 | 2024/09/26 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/02 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-38696 | 2024/09/17 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/01 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-38725 | 2024/10/16 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/01 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia; Loss of consciousness |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|--------------------|--------------------|-------------|--------------------------|--------------------|--------------|-------------------------------|--|---|
| 2954323-2024-38356 | 2024/09/24 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/30 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-38385 | 2024/09/14 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/30 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-38071 | 2024/09/16 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/27 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-38177 | 2024/09/24 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/27 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-37819 | 2024/09/16 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/25 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-37836 | 2024/08/28 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/09/25 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Fatigue; Loss of consciousness |
| 2954323-2024-37452 | 2024/09/12 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/24 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-37556 | 2024/07/28 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/09/24 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | High Readings; Low Readings | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|--------------------|-----------------------|-------------|--------------------------------|-----------------------|--------------|--|--|---|
| 2954323-2024-37602 | 2024/08/27 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/09/24 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia; Loss of consciousness |
| 2954323-2024-37199 | 2024/09/11 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/23 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-37312 | 2024/09/16 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/23 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-37384 | 2024/09/09 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/09/23 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia |
| 2954323-2024-37068 | 2024/09/06 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/20 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-36432 | 2024/08/27 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/18 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-36119 | 2024/08/29 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/17 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|--------------------|-----------------------|-------------|--------------------------------|-----------------------|--------------|--|---|---|
| 2954323-2024-36282 | 2024/08/08 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/09/17 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Detachment of Device or Device Component | Fatigue; Headache; Vomiting; Dizziness; Discomfort; Diabetic Ketoacidosis; Loss of consciousness; Cognitive Changes |
| 2954323-2024-35875 | 2024/09/01 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/09/16 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia |
| 2954323-2024-35555 | 2024/09/09 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/13 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-35464 | 2024/09/05 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/12 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-35298 | 2024/08/14 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/09/11 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Adverse Event Without Identified Device or Use Problem | Pain; Skin Infection |
| 2954323-2024-35302 | 2024/08/29 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/09/11 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | High Readings | Hypoglycemia |
| 2954323-2024-34458 | 2024/08/08 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/09/05 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| 2954323-2024-34164 | 2024/08/26 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/04 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|--------------------|--------------------|-------------|--------------------------|--------------------|--------------|-------------------------------|--|---|
| 2954323-2024-33175 | 2024/08/21 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/08/29 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-32833 | 2024/08/16 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/08/28 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-32980 | 2024/08/12 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/08/28 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-32298 | 2024/08/15 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/08/26 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-31811 | 2024/06/25 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/08/23 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Incorrect, Inadequate or Imprecise Result or Readings | Hypoglycemia; Confusion/ Disorientation |
| 2954323-2024-31398 | 2024/08/10 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/08/21 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-29693 | 2024/07/12 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/08/11 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS SENSOR | Unable to Obtain Readings; Device Displays Incorrect Message | Insufficient Information |
| 2954323-2024-29653 | 2024/07/12 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/08/09 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Detachment of Device or Device Component | Hyperglycemia; Loss of consciousness |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|--------------------|-----------------------|-------------|--------------------------------|-----------------------|--------------|--|--|---|
| 2954323-2024-25655 | 2024/07/13 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/07/17 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-24171 | 2024/06/22 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/07/08 4:00:00 | QBJ | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia; Loss of consciousness; Diaphoresis |
| 2954323-2024-21061 | 2024/05/23 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/06/18 4:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia; Dizziness; Confusion/ Disorientation |
| 2954323-2023-49376 | 2023/10/09 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2023/11/09 5:00:00 | QLG | FREESTYLE LIBRE 2 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia; Loss of consciousness; Diaphoresis |
| 2954323-2024-42946 | 2024/10/07 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/30 4:00:00 | QBJ | FREESTYLE LIBRE 3 PLUS | Adverse Event Without Identified Device or Use Problem | Erythema; Pain; Skin Infection |
| 2954323-2024-42379 | 2024/10/17 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/25 4:00:00 | QBJ | FREESTYLE LIBRE 3 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia; Loss of consciousness; Shaking/Tremors |
| 2954323-2024-41256 | 2024/09/25 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/18 4:00:00 | QBJ | FREESTYLE LIBRE 3 PLUS | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|------------------------|--------------------|-------------|--------------------------|--------------------|--------------|-------------------------------|---|--|
| 2954323-2024-40546 | 2024/09/28 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/10/15 4:00:00 | QBJ | FREESTYLE LIBRE 3 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-40525 | 2024/09/24 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/10/14 4:00:00 | QBJ | FREESTYLE LIBRE 3 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hyperglycemia; Dizziness |
| 2954323-2024-38072 | 2024/09/02 4:00:00 | Malfunction | ABBOTT DIABETES CARE INC | 2024/09/27 4:00:00 | QBJ | FREESTYLE LIBRE 3 PLUS SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| 2954323-2024-33024 | 2024/08/02 4:00:00 | Injury | ABBOTT DIABETES CARE INC | 2024/08/28 4:00:00 | QBJ | FREESTYLE LIBRE 3 PLUS | Unable to Obtain Readings; Device Displays Incorrect Message | Hypoglycemia; Dizziness; Diaphoresis |
| MW5086 448 | 5/2/2019 0:00 | Malfunction | MEDTRONIC , INC. | 5/7/2019 0:00 | MDS | MEDTRONIC GUARDIAN 3 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Low Test Results | Abdominal Pain; Hyperglycemia; Diabetic Ketoacidosis |
| 30098627 00-2024-00899 | 2024/08/05 4:00:00 | Injury | SENSEONICS INC. | 2024/09/10 4:00:00 | QHJ | EVERSENSE E3 SENSOR | Adverse Event Without Identified Device or Use Problem | Skin Infection; Skin Inflammation/ Irritation |
| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
| | | | | | | | | |
| MW51611 55 | 10/6/2024 4:00 | Injury | DEXCOM, INC. | 10/16/2024 4:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|----------------|--------------|--|----------------|---|
| MW5160563 | | Death | DEXCOM, INC. | 10/3/2024 4:00 | QBJ | DEXCOM CONTINUOUS GLUCOSE MONITOR | | Insufficient Information |
| MW5160567 | 9/27/2024 4:00 | Injury | DEXCOM, INC. | 10/3/2024 4:00 | QBJ | DEXCOM G6 SENSOR | | Hyperglycemia; Hypoglycemia |
| MW5160467 | | Injury | DEXCOM, INC. | 10/2/2024 4:00 | QBJ | DEXCOM | | No Clinical Signs, Symptoms or Conditions |
| MW5160472 | | Injury | DEXCOM, INC. | 10/2/2024 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5160333 | | Death | DEXCOM, INC. | 9/30/2024 4:00 | QBJ | DEXCOM | | Insufficient Information |
| MW5160384 | | Malfunction | DEXCOM, INC. | 9/30/2024 4:00 | QBJ | DEXCOM CONTINUOUS GLUCOSE MONITOR SYSTEM | | Rash |
| MW5160420 | | Injury | DEXCOM, INC. | 9/30/2024 4:00 | QBJ | DEXCOM G6 | | Insufficient Information |
| MW5160191 | | Injury | DEXCOM, INC. | 9/26/2024 4:00 | QBJ | DEXCOM CONTINUOUS GLUCOSE MONITOR | | Hyperglycemia |
| MW5160099 | 9/19/2024 4:00 | Malfunction | DEXCOM, INC. | 9/25/2024 4:00 | QBJ | DEXCOM G6 SENSOR | | Insufficient Information |
| MW5160103 | | Malfunction | DEXCOM, INC. | 9/25/2024 4:00 | QBJ | DEXCOM | | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|----------------|-------------|--------------|----------------|--------------|-----------------------------------|---|---|
| MW5160104 | | Malfunction | DEXCOM, INC. | 9/25/2024 4:00 | QBJ | DEXCOM | | No Clinical Signs, Symptoms or Conditions |
| MW5160013 | 9/16/2024 4:00 | Injury | DEXCOM, INC. | 9/24/2024 4:00 | QBJ | DEXCOM SENSOR | | Insufficient Information |
| MW5160023 | 9/17/2024 4:00 | Injury | DEXCOM, INC. | 9/24/2024 4:00 | QBJ | DEXCOM CONTINUOUS GLUCOSE MONITOR | | No Clinical Signs, Symptoms or Conditions |
| MW5159708 | 9/16/2024 4:00 | Injury | DEXCOM, INC. | 9/17/2024 4:00 | QBJ | DEXCOM BLOOD GLUCOSE MONITOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5159446 | 9/5/2024 4:00 | Injury | DEXCOM, INC. | 9/9/2024 4:00 | QBJ | DEXCOM G6 SENSOR | High Test Results | Insufficient Information |
| MW5159455 | 8/2/2024 4:00 | Injury | DEXCOM, INC. | 9/9/2024 4:00 | QBJ | DEXCOM G6 | | Hypoglycemia |
| MW5159311 | 8/31/2024 4:00 | Injury | DEXCOM, INC. | 9/6/2024 4:00 | QBJ | DEXCOM SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5159349 | 9/4/2024 4:00 | Injury | DEXCOM, INC. | 9/6/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5159056 | 8/26/2024 4:00 | Injury | DEXCOM, INC. | 8/29/2024 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5159057 | 8/26/2024 4:00 | Injury | DEXCOM, INC. | 8/29/2024 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-------------------|-------------|--------------|-------------------|--------------|------------------|---|---|
| MW5159058 | 8/26/2024 4:00 | Injury | DEXCOM, INC. | 8/29/2024 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5159081 | 8/26/2024 4:00 | Injury | DEXCOM, INC. | 8/29/2024 4:00 | QBJ | DEXCOM G6 | High Readings; Device Sensing Problem | Hyperglycemia |
| MW5159137 | 8/27/2024 4:00 | Malfunction | DEXCOM, INC. | 8/29/2024 4:00 | QBJ | DEXCOM G6 SENSOR | | Insufficient Information |
| MW5158710 | | Injury | DEXCOM, INC. | 8/21/2024 4:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5158575 | 8/14/2024 4:00 | Injury | DEXCOM, INC. | 8/16/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5158302 | | Injury | DEXCOM, INC. | 8/9/2024 4:00 | QBJ | DEXCOM | | Insufficient Information |
| MW5158257 | 8/5/2024 4:00 | Injury | DEXCOM, INC. | 8/7/2024 4:00 | QBJ | DEXCOM | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5157864 | | Injury | DEXCOM, INC. | 7/30/2024 4:00 | QBJ | DEXCOM | | Hypoglycemia |
| MW5157737 | 7/22/2024 4:00 | Injury | DEXCOM, INC. | 7/26/2024 4:00 | QBJ | DEXCOM G6 SENSOR | | Discomfort |
| MW5156911 | 6/26/2024 4:00 | Malfunction | DEXCOM, INC. | 7/1/2024 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-------------------|-------------|--------------|-------------------|--------------|------------------|---|---|
| MW5156505 | 6/17/2024 4:00 | Injury | DEXCOM, INC. | 6/21/2024 4:00 | QBJ | DEXCOM G6 | | Insufficient Information; No Clinical Signs, Symptoms or Conditions |
| MW5156459 | | Injury | DEXCOM, INC. | 6/20/2024 4:00 | QBJ | DEXCOM | | Insufficient Information |
| MW5156471 | | Injury | DEXCOM, INC. | 6/20/2024 4:00 | QBJ | DEXCOM | | Insufficient Information |
| MW5155988 | | Injury | DEXCOM, INC. | 6/7/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |
| MW5155768 | 6/3/2024 4:00 | Malfunction | DEXCOM, INC. | 6/4/2024 4:00 | QBJ | DEXCOM G6 | | Insufficient Information |
| MW5153878 | 4/11/2024 4:00 | Malfunction | DEXCOM, INC. | 4/15/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |
| MW5153513 | 4/3/2024 4:00 | Injury | DEXCOM, INC. | 4/3/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |
| MW5153215 | 3/23/2024 4:00 | Injury | DEXCOM, INC. | 3/26/2024 4:00 | QBJ | DEXCOM SENSOR | High Readings | No Clinical Signs, Symptoms or Conditions |
| MW5153027 | 3/18/2024 4:00 | Injury | DEXCOM, INC. | 3/20/2024 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hyperglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-------------------|-------------|--------------|-------------------|--------------|---|---|---|
| MW5153005 | 2/20/2024 5:00 | Malfunction | DEXCOM, INC. | 3/19/2024 4:00 | QBJ | DEXCOM G6 CONTINUEO US GLUCOSE MONITOR SYSTEM | | Foreign Body In Patient |
| MW5152444 | | Injury | DEXCOM, INC. | 3/6/2024 5:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5151801 | 2/11/2024 5:00 | Injury | DEXCOM, INC. | 2/21/2024 5:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5151825 | | Injury | DEXCOM, INC. | 2/21/2024 5:00 | QBJ | DEXCOM | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5151167 | 2/5/2024 5:00 | Injury | DEXCOM, INC. | 2/7/2024 5:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5150660 | 1/20/2024 5:00 | Injury | DEXCOM, INC. | 1/23/2024 5:00 | QBJ | DEXCOM G6 SENSOR | High Readings | No Clinical Signs, Symptoms or Conditions |
| MW5150681 | 1/21/2024 5:00 | Malfunction | DEXCOM, INC. | 1/23/2024 5:00 | QBJ | DEXCOM G6 TRANSMITTE R | | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-----------------|-------------|--------------|-----------------|--------------|----------------------|---|---|
| MW5150490 | | Injury | DEXCOM, INC. | 1/17/2024 5:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings | Insufficient Information |
| MW5150101 | | Injury | DEXCOM, INC. | 1/9/2024 5:00 | QBJ | DEXCOM G6 | High Readings; Device Sensing Problem | Hyperglycemia |
| MW5149414 | | Injury | DEXCOM, INC. | 12/19/2023 5:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5148679 | | Injury | DEXCOM, INC. | 12/1/2023 5:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | No Clinical Signs, Symptoms or Conditions |
| MW5148559 | | Injury | DEXCOM, INC. | 11/29/2023 5:00 | QBJ | DEXCOM G6 | Incorrect, Inadequate or Imprecise Result or Readings; Failure to Sense | Hyperglycemia |
| MW5148569 | 11/23/2023 5:00 | Injury | DEXCOM, INC. | 11/29/2023 5:00 | QBJ | DEXCOM G6 CGM SENSOR | | Unspecified Infection; Rash |
| MW5148448 | | Injury | DEXCOM, INC. | 11/22/2023 5:00 | QBJ | DEXCOM G6 | High Readings; Calibration Problem | Insufficient Information |
| MW5148455 | 11/18/2023 5:00 | Malfunction | DEXCOM, INC. | 11/22/2023 5:00 | QBJ | DEXCOM G6 CGM SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5148368 | | Injury | DEXCOM, INC. | 11/21/2023 5:00 | QBJ | DEXCOM G6 | | Hyperglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-----------------|-------------|--------------|-----------------|--------------|-------------------------------------|---|---|
| MW5148145 | | Injury | DEXCOM, INC. | 11/14/2023 5:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5148148 | | Injury | DEXCOM, INC. | 11/14/2023 5:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5148101 | | Injury | DEXCOM, INC. | 11/13/2023 5:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | No Clinical Signs, Symptoms or Conditions |
| MW5148022 | | Injury | DEXCOM, INC. | 11/9/2023 5:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5147844 | | Injury | DEXCOM, INC. | 11/6/2023 5:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings | Hypoglycemia |
| MW5147801 | | Injury | DEXCOM, INC. | 11/3/2023 4:00 | QBJ | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem | Hyperglycemia |
| MW5147758 | | Injury | DEXCOM, INC. | 11/2/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5147654 | | Injury | DEXCOM, INC. | 11/1/2023 4:00 | QBJ | DEXCOM G6 | | Insufficient Information |
| MW5147085 | 10/12/2023 4:00 | Malfunction | DEXCOM, INC. | 10/18/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE SENSOR | | Insufficient Information |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|--------------------|-------------|--------------|--------------------|--------------|-------------------------------------|------------------|--|
| MW5147086 | 10/12/2023 4:00 | Malfunction | DEXCOM, INC. | 10/18/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE SENSOR | | Insufficient Information |
| MW5147087 | 10/12/2023 4:00 | Malfunction | DEXCOM, INC. | 10/18/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE SENSOR | Defective Device | Insufficient Information |
| MW5146930 | 10/8/2023 4:00 | Injury | DEXCOM, INC. | 10/12/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia; Hypoglycemia; Diabetic Ketoacidosis |
| MW5146931 | | Injury | DEXCOM, INC. | 10/12/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5146871 | | Injury | DEXCOM, INC. | 10/11/2023 4:00 | QBJ | DEXCOM G6 | | Hypoglycemia |
| MW5146505 | | Injury | DEXCOM, INC. | 10/3/2023 4:00 | QBJ | DEXCOM G6 | | Hypoglycemia |
| MW5146340 | | Injury | DEXCOM, INC. | 9/29/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5146355 | | Injury | DEXCOM, INC. | 9/29/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5146362 | | Injury | DEXCOM, INC. | 9/29/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5146370 | | Injury | DEXCOM, INC. | 9/29/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-------------------|-------------|--------------|-------------------|--------------|--|----------------|---|
| MW5146223 | 9/25/2023 4:00 | Malfunction | DEXCOM, INC. | 9/27/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5146224 | 9/25/2023 4:00 | Malfunction | DEXCOM, INC. | 9/27/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR | | No Clinical Signs, Symptoms or Conditions |
| MW5146225 | 9/25/2023 4:00 | Malfunction | DEXCOM, INC. | 9/27/2023 4:00 | QBJ | DEXCOM G6 CONTINUOUS GLUCOSE MONITOR TRANSMITTER | | No Clinical Signs, Symptoms or Conditions |
| MW5146203 | | Injury | DEXCOM, INC. | 9/26/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5146144 | 9/20/2023 4:00 | Injury | DEXCOM, INC. | 9/25/2023 4:00 | QBJ | DEXCOM G6 | | Insufficient Information |
| MW5146150 | | Injury | DEXCOM, INC. | 9/25/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5146110 | | Injury | DEXCOM, INC. | 9/22/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5145980 | | Injury | DEXCOM, INC. | 9/20/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5145922 | | Injury | DEXCOM, INC. | 9/19/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|------------------|------------|--------------|-------------------|--------------|------------|----------------|---|
| MW5145630 | 9/9/2023 4:00 | Injury | DEXCOM, INC. | 9/13/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5145634 | 9/9/2023 4:00 | Injury | DEXCOM, INC. | 9/13/2023 4:00 | QBJ | DEXCOM | | Skin Inflammation/ Irritation |
| MW5145381 | | Injury | DEXCOM, INC. | 9/7/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5145386 | | Injury | DEXCOM, INC. | 9/7/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5145275 | | Injury | DEXCOM, INC. | 9/6/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW5144951 | | Injury | DEXCOM, INC. | 8/24/2023 4:00 | QBJ | DEXCOM G6 | | Hypoglycemia |
| MW5144952 | | Injury | DEXCOM, INC. | 8/24/2023 4:00 | QBJ | DEXCOM G6 | | Hypoglycemia |
| MW5144652 | | Injury | DEXCOM, INC. | 8/15/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5144656 | | Injury | DEXCOM, INC. | 8/15/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW5144657 | | Injury | DEXCOM, INC. | 8/15/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |

| Report Number | Event Date | Event Type | Manufacturer | Date Received | Product Code | Brand Name | Device Problem | Patient Problem |
|---------------|-------------------|------------|-----------------|-------------------|--------------|---------------------|---|--|
| MW51228 81 | 7/25/2023 4:00 | Injury | DEXCOM, INC. | 8/9/2023 4:00 | QBJ | DEXCOM G6 | | Dehydration; Diarrhea; Fatigue; Hyperglycemia; Vomiting; Diabetic Ketoacidosis |
| MW51206 29 | | Injury | DEXCOM, INC. | 8/4/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW51203 50 | | Injury | DEXCOM, INC. | 7/28/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW51202 17 | | Injury | DEXCOM, INC. | 7/26/2023 4:00 | QBJ | DEXCOM G6 | | Hyperglycemia |
| MW51199 83 | | Injury | DEXCOM, INC. | 7/21/2023 4:00 | QBJ | DEXCOM G6 | | No Clinical Signs, Symptoms or Conditions |
| MW51192 29 | 7/5/2023 4:00 | Injury | DEXCOM, INC. | 7/7/2023 4:00 | QBJ | DEXCOM G6 SENSOR | Pumping Stopped; Low Readings | Hyperglycemia |
| MW51188 23 | 6/21/2023 4:00 | Injury | DEXCOM, INC. | 6/23/2023 4:00 | MDS | DEXCOM G6 SENSOR | Incorrect, Inadequate or Imprecise Result or Readings; Therapeutic or Diagnostic Output Failure | Hypoglycemia |

Abbreviations.